



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**Metals**

**- 2b -**

**CRDL STANDARD FOR AA & ICP**

**Client:** G Environmental

**SDG No.:** Q3058

**Contract:** GENV01

**Lab Code:** ACE

**Initial Calibration Source:** \_\_\_\_\_

**Continuing Calibration Source:** \_\_\_\_\_

| Sample ID | Analyte   | Result<br>ug/L | True Value<br>ug/L | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|-----------|----------------|--------------------|---------------|---------------------------|---|------------------|------------------|---------------|
| CRI01     | Iron      | 92.9           | 100                | 93            | 65 - 135                  | P | 09/12/2025       | 13:47            | LB137179      |
|           | Manganese | 19.5           | 20.0               | 98            | 65 - 135                  | P | 09/12/2025       | 13:47            | LB137179      |
|           | Sodium    | 1870           | 2000               | 94            | 65 - 135                  | P | 09/12/2025       | 13:47            | LB137179      |
| CRI01     | Iron      | 104            | 100                | 104           | 65 - 135                  | P | 09/15/2025       | 11:47            | LB137198      |
|           | Manganese | 19.6           | 20.0               | 98            | 65 - 135                  | P | 09/15/2025       | 11:47            | LB137198      |
|           | Sodium    | 1840           | 2000               | 92            | 65 - 135                  | P | 09/15/2025       | 11:47            | LB137198      |