

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental

SDG No.: Q3058

Contract: GENV01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169680BS							
Iron	ug/L	1500	1480		99	80 - 120	P
Manganese	ug/L	100	100		100	80 - 120	P
Sodium	ug/L	1500	1490		99	80 - 120	P