



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091025\
Data File : VY023320.D
Acq On : 10 Sep 2025 08:38
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090825S.M
Title : SW846 8260
Last Update : Wed Sep 10 01:44:33 2025



AutoFind: Scans 1752, 1753, 1754; Background Corrected with Scan 1744

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.4	24923	PASS
75	95	30	60	49.8	71453	PASS
95	95	100	100	100.0	143568	PASS
96	95	5	9	6.7	9647	PASS
173	174	0.00	2	1.1	1481	PASS
174	95	50	100	93.1	133704	PASS
175	174	5	9	7.4	9879	PASS
176	174	95	101	95.1	127181	PASS
177	176	5	9	6.6	8360	PASS