

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN082125

Review By	John Carlone	Review On	8/22/2025 9:17:52 AM
Supervise By	Maresh Dadoda	Supervise On	8/25/2025 8:11:38 AM
SubDirectory	VN082125	HP Acquire Method	HP Processing Method 82N082125W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135214 VP135215,VP135216,VP135217,VP135218,VP135219,VP135220 VP135221

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087614.D	21 Aug 2025 09:30		JC\MD	Ok
2	VSTDCCC020	VSTDCCC020	VN087615.D	21 Aug 2025 10:04	Need ICAL	JC\MD	Not Ok
3	VN0821WBS01	VN0821WBS01	VN087616.D	21 Aug 2025 10:37		JC\MD	Not Ok
4	VN0821WBSD01	VN0821WBSD01	VN087617.D	21 Aug 2025 11:11		JC\MD	Not Ok
5	VN0821WBL01	VN0821WBL01	VN087618.D	21 Aug 2025 11:36		JC\MD	Not Ok
6	VSTDICC001	VSTDICC001	VN087619.D	21 Aug 2025 12:09	LR-10,20	JC\MD	Ok,M
7	VSTDICC005	VSTDICC005	VN087620.D	21 Aug 2025 13:08		JC\MD	Ok,M
8	VSTDICC020	VSTDICC020	VN087621.D	21 Aug 2025 13:41	method not used for #N-amyle Acetate	JC\MD	Ok,M
9	VSTDICCC050	VSTDICCC050	VN087622.D	21 Aug 2025 14:02		JC\MD	Ok,M
10	VSTDICC100	VSTDICC100	VN087623.D	21 Aug 2025 14:23		JC\MD	Ok,M
11	VSTDICC150	VSTDICC150	VN087624.D	21 Aug 2025 14:44		JC\MD	Ok,M
12	IBLK	IBLK	VN087625.D	21 Aug 2025 15:05		JC\MD	Ok
13	VSTDICV050	ICVVN082125	VN087626.D	21 Aug 2025 15:47	Icv Failed for Com.# 64 for DOD	JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN091525

Review By	Mahesh Dadoda	Review On	9/16/2025 7:59:53 AM
Supervise By	Semsettin Yesilyurt	Supervise On	9/16/2025 8:03:28 AM
SubDirectory	VN091525	HP Acquire Method	HP Processing Method 82N082125W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135447
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135448,VP135449

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087819.D	15 Sep 2025 08:18		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN087820.D	15 Sep 2025 08:39	pH#Lot#V12668	JC\MD	Ok,M
3	VN0915MBL01	VN0915MBL01	VN087821.D	15 Sep 2025 09:14		JC\MD	Ok
4	VN0915WBL01	VN0915WBL01	VN087822.D	15 Sep 2025 09:35		JC\MD	Ok
5	VN0915WBS01	VN0915WBS01	VN087823.D	15 Sep 2025 09:56		JC\MD	Ok,M
6	VN0915WBSD01	VN0915WBSD01	VN087824.D	15 Sep 2025 10:29		JC\MD	Ok,M
7	Q3083-01	MW4S	VN087825.D	15 Sep 2025 10:50	vial B pH<2	JC\MD	Ok
8	Q3088-02	90525-LQ	VN087826.D	15 Sep 2025 11:11	vial A pH<2 brown/ turbid sample	JC\MD	Ok
9	PB169678TB	PB169678TB	VN087827.D	15 Sep 2025 11:32		JC\MD	Ok
10	Q3082-02	WC-29	VN087828.D	15 Sep 2025 11:53	vial A pH#5.0	JC\MD	Ok
11	Q3082-04	WC-30	VN087829.D	15 Sep 2025 12:14	vial A pH#5.0	JC\MD	Ok,M
12	Q3082-06	WC-31	VN087830.D	15 Sep 2025 12:35	vial A pH#5.0	JC\MD	Ok
13	Q3082-08	WC-32	VN087831.D	15 Sep 2025 12:56	vial A pH#5.0	JC\MD	Ok
14	Q3082-10	WC-33	VN087832.D	15 Sep 2025 13:17	vial A pH#5.0	JC\MD	Ok,M
15	Q3092-01	291376	VN087833.D	15 Sep 2025 13:38	vial A pH#5.0	JC\MD	Ok
16	Q3092-02	291376	VN087834.D	15 Sep 2025 13:59	vial A pH#5.0	JC\MD	Ok
17	Q3092-03	279182	VN087835.D	15 Sep 2025 14:20	vial A pH#5.0	JC\MD	Ok,M

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN091525

Review By	Maresh Dadoda	Review On	9/16/2025 7:59:53 AM
Supervise By	Semsettin Yesilyurt	Supervise On	9/16/2025 8:03:28 AM
SubDirectory	VN091525	HP Acquire Method	HP Processing Method 82N082125W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135447		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135448,VP135449		

18	Q3092-04	279182	VN087836.D	15 Sep 2025 14:41	vial A pH#5.0	JC\MD	Ok
19	VN0915MBS01	VN0915MBS01	VN087837.D	15 Sep 2025 15:02		JC\MD	Ok,M
20	VN0915MBSD01	VN0915MBSD01	VN087838.D	15 Sep 2025 15:22		JC\MD	Ok,M
21	Q3058-01ME	RPXY5ME	VN087839.D	15 Sep 2025 15:43		JC\MD	Ok
22	Q3075-04	MOO-25-0258	VN087840.D	15 Sep 2025 16:04	cloth material	JC\MD	Ok
23	Q3074-01	MOO-25-0141	VN087841.D	15 Sep 2025 16:25	cloth material	JC\MD	Ok
24	Q3073-02	BEL-25-0023	VN087842.D	15 Sep 2025 16:46	cloth material	JC\MD	Ok
25	Q3075-06	MOO-25-0259	VN087843.D	15 Sep 2025 17:06	cloth material	JC\MD	Ok
26	Q3077-01	LAW-25-0137	VN087844.D	15 Sep 2025 17:27	cloth material run straight MEOH	JC\MD	Not Ok
27	VSTDCCC050	VSTDCCC050EC	VN087845.D	15 Sep 2025 18:09		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY090825

Review By	Maresh Dadoda	Review On	9/9/2025 3:13:06 PM
Supervise By	Semsettin Yesilyurt	Supervise On	9/19/2025 2:10:25 AM
SubDirectory	VY090825	HP Acquire Method	MSVOA_Y HP Processing Method 82y090825s.m

STD. NAME	STD REF.#
Tune/Reschk	VP135392
Initial Calibration Stds	VP135393,VP135394,VP135395,VP135396,VP135397,VP135398
CCC	
Internal Standard/PEM	VP133934
ICV/I.BLK	VP135399
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023302.D	08 Sep 2025 09:07		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY023303.D	08 Sep 2025 10:00		SY/MD	Ok,M
3	VSTDICC010	VSTDICC010	VY023304.D	08 Sep 2025 10:23		SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY023305.D	08 Sep 2025 11:01		SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY023306.D	08 Sep 2025 11:23		SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VY023307.D	08 Sep 2025 11:46		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VY023308.D	08 Sep 2025 12:08		SY/MD	Ok,M
8	VIBLK	VIBLK	VY023309.D	08 Sep 2025 12:32		SY/MD	Ok
9	VSTDICV050	ICVVY090825	VY023310.D	08 Sep 2025 12:55		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY091025

Review By	Semsettin Yesilyurt	Review On	9/11/2025 5:40:36 PM
Supervise By	Maresh Dadoda	Supervise On	9/19/2025 2:06:55 AM
SubDirectory	VY091025	HP Acquire Method	MSVOA_Y HP Processing Method 82y090825s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135409
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135410,VP135411 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023320.D	10 Sep 2025 08:38		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY023321.D	10 Sep 2025 09:09		SY/MD	Ok,M
3	VY0910SBL01	VY0910SBL01	VY023322.D	10 Sep 2025 09:39		SY/MD	Ok
4	VY0910SBS01	VY0910SBS01	VY023323.D	10 Sep 2025 10:11		SY/MD	Ok,M
5	VY0910SBSD01	VY0910SBSD01	VY023324.D	10 Sep 2025 10:33		SY/MD	Ok,M
6	Q3055-01	OR-03-090925	VY023325.D	10 Sep 2025 11:39	vial-A	SY/MD	Ok
7	Q3056-01	OK-02-090925	VY023326.D	10 Sep 2025 12:02	vial-A	SY/MD	Ok
8	Q3059-01	TR-06-09092025	VY023327.D	10 Sep 2025 12:25	vial-A	SY/MD	Ok
9	Q3058-01	RPXY5	VY023328.D	10 Sep 2025 12:49	vial-A Need MeOH	SY/MD	Dilution
10	Q3058-02	RPXY4	VY023329.D	10 Sep 2025 13:12	vial-A	SY/MD	Ok
11	Q3066-01	354	VY023330.D	10 Sep 2025 14:02	vial-A	SY/MD	Ok
12	Q3066-03	VNJ-242	VY023331.D	10 Sep 2025 14:26	vial-A	SY/MD	Ok
13	Q3065-01	VNJ-204	VY023332.D	10 Sep 2025 14:49	vial-A Internal Standard fail	SY/MD	ReRun
14	VIBLK	VIBLK	VY023333.D	10 Sep 2025 15:13		SY/MD	Ok
15	VSTDCCC050	VSTDCCC050EC	VY023334.D	10 Sep 2025 15:35		SY/MD	Ok,M

M : Manual Integration