

Surrogate Summary

SDG No.: **Q3058**

Client: **G Environmental**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3058-01	RPXY5	1,2-Dichloroethane-d4	50	62.0	124		70 (63)	130 (155)
		Dibromofluoromethane	50	54.2	108		70 (70)	130 (134)
		Toluene-d8	50	52.0	104		70 (74)	130 (123)
		4-Bromofluorobenzene	50	54.6	109		70 (17)	130 (146)
Q3058-01ME	RPXY5ME	1,2-Dichloroethane-d4	50	50.6	101		70 (63)	130 (155)
		Dibromofluoromethane	50	50.6	101		70 (70)	130 (134)
		Toluene-d8	50	46.8	94		70 (74)	130 (123)
		4-Bromofluorobenzene	50	39.8	80		70 (17)	130 (146)
Q3058-02	RPXY4	1,2-Dichloroethane-d4	50	67.9	136	*	70 (63)	130 (155)
		Dibromofluoromethane	50	57.9	116		70 (70)	130 (134)
		Toluene-d8	50	50.9	102		70 (74)	130 (123)
		4-Bromofluorobenzene	50	55.5	111		70 (17)	130 (146)
VN0915MBL01	VN0915MBL01	1,2-Dichloroethane-d4	50	49.9	100		70 (63)	130 (155)
		Dibromofluoromethane	50	52.3	105		70 (70)	130 (134)
		Toluene-d8	50	48.3	97		70 (74)	130 (123)
		4-Bromofluorobenzene	50	45.6	91		70 (17)	130 (146)
VN0915MBS01	VN0915MBS01	1,2-Dichloroethane-d4	50	48.5	97		70 (63)	130 (155)
		Dibromofluoromethane	50	49.4	99		70 (70)	130 (134)
		Toluene-d8	50	45.6	91		70 (74)	130 (123)
		4-Bromofluorobenzene	50	46.4	93		70 (17)	130 (146)
VY0910SBL01	VY0910SBL01	1,2-Dichloroethane-d4	50	61.2	122		70 (63)	130 (155)
		Dibromofluoromethane	50	53.7	107		70 (70)	130 (134)
		Toluene-d8	50	51.7	103		70 (74)	130 (123)
		4-Bromofluorobenzene	50	49.4	99		70 (17)	130 (146)
VY0910SBS01	VY0910SBS01	1,2-Dichloroethane-d4	50	49.6	99		70 (63)	130 (155)
		Dibromofluoromethane	50	49.3	99		70 (70)	130 (134)
		Toluene-d8	50	50.5	101		70 (74)	130 (123)
		4-Bromofluorobenzene	50	50.6	101		70 (17)	130 (146)
VY0910SBSD01	VY0910SBSD01	1,2-Dichloroethane-d4	50	48.3	97		70 (63)	130 (155)
		Dibromofluoromethane	50	48.5	97		70 (70)	130 (134)
		Toluene-d8	50	49.0	98		70 (74)	130 (123)
		4-Bromofluorobenzene	50	48.7	97		70 (17)	130 (146)

