

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q3058
 Lab File ID: VN087820.D Date Analyzed: 09/15/2025
 Instrument ID: MSVOA_N Time Analyzed: 08:39
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	259703	13.765				
UPPER LIMIT	519406	14.265				
LOWER LIMIT	129852	13.265				
EPA SAMPLE NO.						
RPXY5ME	170499	13.77				
VN0915MBL01	193776	13.77				
VN0915MBS01	213436	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.