

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q3058
 Lab File ID: VY023321.D Date Analyzed: 09/10/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:09
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	341214	13.346				
UPPER LIMIT	682428	13.846				
LOWER LIMIT	170607	12.846				
EPA SAMPLE NO.						
RPXY5	418953	13.35				
RPXY4	442394	13.35				
VY0910SBL01	415042	13.35				
VY0910SBS01	350574	13.35				
VY0910SBSD01	341290	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.