

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG NO.: Q3058

Lab File ID: VN087614.D

BFB Injection Date: 08/21/2025

Instrument ID: MSVOA_N

BFB Injection Time: 09:30

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	25.8
75	30.0 - 60.0% of mass 95	58.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.3
173	Less than 2.0% of mass 174	1.1 (1.6) 1
174	50.0 - 100.0% of mass 95	69
175	5.0 - 9.0% of mass 174	5.6 (8.1) 1
176	95.0 - 101.0% of mass 174	67.7 (98.1) 1
177	5.0 - 9.0% of mass 176	4.8 (7.1) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VN087619.D	08/21/2025	12:09
VSTDIC005	VSTDIC005	VN087620.D	08/21/2025	13:08
VSTDIC020	VSTDIC020	VN087621.D	08/21/2025	13:41
VSTDIC050	VSTDIC050	VN087622.D	08/21/2025	14:02
VSTDIC100	VSTDIC100	VN087623.D	08/21/2025	14:23
VSTDIC150	VSTDIC150	VN087624.D	08/21/2025	14:44