



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091025\
Data File : VN087773.D
Acq On : 10 Sep 2025 11:13
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Title : SW846 8260
Last Update : Fri Aug 22 02:55:42 2025



AutoFind: Scans 1848, 1849, 1850; Background Corrected with Scan 1840

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.0	10114	PASS
75	95	30	60	56.3	25907	PASS
95	95	100	100	100.0	46035	PASS
96	95	5	9	7.4	3409	PASS
173	174	0.00	2	0.7	236	PASS
174	95	50	100	72.4	33341	PASS
175	174	5	9	8.1	2709	PASS
176	174	95	101	95.5	31851	PASS
177	176	5	9	7.1	2248	PASS

Average of 12.823 to 12.835 min.: VN087773.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
35.95	552	50.00	10114	60.05	327	71.60	61
37.00	2621	51.00	3078	61.00	2848	71.90	35
38.05	3091	51.70	90	62.05	2423	72.95	244
39.00	1641	52.10	52	62.95	2583	74.05	995
40.90	60	52.80	70	63.85	595	75.00	25907
43.95	630	54.90	199	64.95	597	75.95	2017
45.00	625	55.90	639	66.75	119	76.80	107
45.90	172	56.10	301	67.10	84	77.00	247
47.00	875	57.00	1541	67.95	5434	77.95	321
47.95	412	57.90	80	68.90	5780	78.90	2372
48.95	2312	59.85	394	69.95	681	79.80	341

Average of 12.823 to 12.835 min.: VN087773.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
79.95	533	95.00	46035	116.70	171	141.90	64
80.85	2318	95.95	3409	116.95	167	142.85	884
81.85	548	103.55	146	117.90	141	144.50	56
82.60	55	103.90	339	118.10	77	144.80	149
85.80	140	105.95	237	118.90	470	147.60	74
86.90	2213	109.80	53	127.60	71	149.60	60
87.95	2246	111.90	106	127.90	104	156.50	106
90.90	212	112.75	118	128.50	60	157.00	59
91.95	1773	114.90	79	134.70	166	170.25	115
92.95	2919	115.80	83	136.70	65	171.85	751
94.00	6994	116.10	91	140.85	767	172.50	295

Average of 12.823 to 12.835 min.: VN087773.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
173.00	236						
173.90	33341						
174.95	2709						
175.90	31851						
176.95	2248						
177.70	105						
192.80	87						
206.80	219						