

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
Lab Code: ACE SDG NO.: Q3058
Lab File ID: VN087774.D Date Analyzed: 09/10/2025
Instrument ID: MSVOA_N Time Analyzed: 11:47
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	241352	8.21	467783	9.08	458675	11.84
UPPER LIMIT	482704	8.706	935566	9.582	917350	12.341
LOWER LIMIT	120676	7.706	233892	8.582	229338	11.341
EPA SAMPLE NO.						
MW4	196528	8.21	453885	9.08	434796	11.84
VN0910WBL01	177881	8.21	413161	9.08	392580	11.84
VN0910WBS01	247948	8.21	481179	9.08	443678	11.85
VN0910WBSD01	258583	8.21	480224	9.08	451507	11.84

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.