

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q3058
 Lab File ID: VN087774.D Date Analyzed: 09/10/2025
 Instrument ID: MSVOA_N Time Analyzed: 11:47
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	238791	13.77				
UPPER LIMIT	477582	14.27				
LOWER LIMIT	119396	13.27				
EPA SAMPLE NO.						
MW4	206220	13.77				
VN0910WBL01	179388	13.77				
VN0910WBS01	236791	13.77				
VN0910WBSD01	236702	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.