

## Matrix Spike Summary

|                   |                          |                                         |          |
|-------------------|--------------------------|-----------------------------------------|----------|
| <b>Client:</b>    | CHA Companies, Inc.      | <b>SDG No.:</b>                         | Q3061    |
| <b>Project:</b>   | NYSOMH Emergency Service | <b>Sample ID:</b>                       | Q3061-01 |
| <b>Client ID:</b> | B2-W16-RM4080-SINKMS     | <b>Percent Solids for Spike Sample:</b> | 0        |

| Analyte           | Units | Acceptance<br>Limit %R | Spiked<br>Result | Conc.<br>Qualifier | Sample<br>Result | Conc.<br>Qualifier | Spike<br>Added | Dilution<br>Factor | %<br>Rec | Qual | Analysis<br>Date |
|-------------------|-------|------------------------|------------------|--------------------|------------------|--------------------|----------------|--------------------|----------|------|------------------|
| Residual Chlorine | mg/L  | 71-148                 | 0.44             |                    | 0.093            | J                  | 0.4            | 1                  | 86       |      | 09/10/2025       |

### Matrix Spike Summary

|                   |                          |                                         |          |
|-------------------|--------------------------|-----------------------------------------|----------|
| <b>Client:</b>    | CHA Companies, Inc.      | <b>SDG No.:</b>                         | Q3061    |
| <b>Project:</b>   | NYSOMH Emergency Service | <b>Sample ID:</b>                       | Q3061-01 |
| <b>Client ID:</b> | B2-W16-RM4080-SINKMSD    | <b>Percent Solids for Spike Sample:</b> | 0        |

| Analyte           | Units | Acceptance<br>Limit %R | Spiked<br>Result | Conc.<br>Qualifier | Sample<br>Result | Conc.<br>Qualifier | Spike<br>Added | Dilution<br>Factor | %<br>Rec | Qual | Analysis<br>Date |
|-------------------|-------|------------------------|------------------|--------------------|------------------|--------------------|----------------|--------------------|----------|------|------------------|
| Residual Chlorine | mg/L  | 71-148                 | 0.43             |                    | 0.093            | J                  | 0.4            | 1                  | 84       |      | 09/10/2025       |