



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Elegant Jewelers Mfg. Co. Inc.

Project Name: Waste Water 2025

Project # N/A

Order ID # Q3063

Test Name: Cyanide,Cyanide-Amenable

A. Number of Samples and Date of Receipt:

1 Water sample was received on 09/11/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Cyanide,Cyanide-Amenable. This data package contains results for Cyanide,Cyanide-Amenable.

C. Analytical Techniques:

The analysis of Cyanide-Amenable was based on method SM4500-CN B,G Cyanide-Amenable and The analysis of Cyanide was based on method SM4500-CN C,E.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all compounds.

The Duplicate analysis met criteria for all compounds.

The Matrix Spike analysis met criteria for all compounds.

The Matrix Spike Duplicate analysis met criteria for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

E. Additional Comments:

Samples were received on 09/11/2025 09:50 and composited in the Lab on 09/11/2025 10:00.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____