

Initial and Continuing Calibration Verification

Client: Elegant Jewelers Mfg. Co. Inc.

SDG No.: Q3063

Project: Waste Water 2025

RunNo.: LB137201

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Cyanide	mg/L	0.096	0.099	97	85-115	09/16/2025
Sample ID: CCV1 Cyanide	mg/L	0.24	0.25	96	90-110	09/16/2025
Sample ID: CCV2 Cyanide	mg/L	0.25	0.25	100	90-110	09/16/2025

Initial and Continuing Calibration Verification

Client: Elegant Jewelers Mfg. Co. Inc.

SDG No.: Q3063

Project: Waste Water 2025

RunNo.: LB137201

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------