



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Elegant Jewelers Mfg. Co. Inc.

SDG No.: Q3063

Contract: ELEG01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB28	Mercury	0.027	+/-0.2	U	0.20	CV	09/15/2025	12:07	LB137189

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Elegant Jewelers Mfg. Co. Inc.

SDG No.: Q3063

Contract: ELEG01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB85	Mercury	0.027	+/-0.2	U	0.20	CV	09/15/2025	12:11	LB137189
CCB86	Mercury	0.027	+/-0.2	U	0.20	CV	09/15/2025	13:10	LB137189
CCB87	Mercury	0.027	+/-0.2	U	0.20	CV	09/15/2025	13:21	LB137189

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Elegant Jewelers Mfg. Co. Inc.

SDG No.: Q3063

Contract: ELEG01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cadmium	0.50	+/-3	U	6.00	P	09/16/2025	11:33	LB137209
	Chromium	2.12	+/-5	U	10.0	P	09/16/2025	11:33	LB137209
	Copper	4.60	+/-10	U	20.0	P	09/16/2025	11:33	LB137209
	Lead	2.30	+/-6	U	12.0	P	09/16/2025	11:33	LB137209
	Molybdenum	18.7	+/-100	U	200	P	09/16/2025	11:33	LB137209
	Nickel	3.06	+/-20	U	40.0	P	09/16/2025	11:33	LB137209
	Silver	1.62	+/-5	U	10.0	P	09/16/2025	11:33	LB137209
	Zinc	16.7	+/-20	U	40.0	P	09/16/2025	11:33	LB137209

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Elegant Jewelers Mfg. Co. Inc.

SDG No.: Q3063

Contract: ELEG01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cadmium	0.50	+/-3	U	6.00	P	09/16/2025	12:05	LB137209
	Chromium	2.12	+/-5	U	10.0	P	09/16/2025	12:05	LB137209
	Copper	4.60	+/-10	U	20.0	P	09/16/2025	12:05	LB137209
	Lead	2.30	+/-6	U	12.0	P	09/16/2025	12:05	LB137209
	Molybdenum	18.7	+/-100	U	200	P	09/16/2025	12:05	LB137209
	Nickel	3.06	+/-20	U	40.0	P	09/16/2025	12:05	LB137209
	Silver	1.62	+/-5	U	10.0	P	09/16/2025	12:05	LB137209
	Zinc	16.7	+/-20	U	40.0	P	09/16/2025	12:05	LB137209
CCB02	Cadmium	0.50	+/-3	U	6.00	P	09/16/2025	13:43	LB137209
	Chromium	2.12	+/-5	U	10.0	P	09/16/2025	13:43	LB137209
	Copper	4.60	+/-10	U	20.0	P	09/16/2025	13:43	LB137209
	Lead	2.30	+/-6	U	12.0	P	09/16/2025	13:43	LB137209
	Molybdenum	18.7	+/-100	U	200	P	09/16/2025	13:43	LB137209
	Nickel	3.06	+/-20	U	40.0	P	09/16/2025	13:43	LB137209
	Silver	1.62	+/-5	U	10.0	P	09/16/2025	13:43	LB137209
	Zinc	16.7	+/-20	U	40.0	P	09/16/2025	13:43	LB137209
CCB03	Cadmium	0.50	+/-3	U	6.00	P	09/16/2025	14:47	LB137209
	Chromium	2.12	+/-5	U	10.0	P	09/16/2025	14:47	LB137209
	Copper	4.60	+/-10	U	20.0	P	09/16/2025	14:47	LB137209
	Lead	2.30	+/-6	U	12.0	P	09/16/2025	14:47	LB137209
	Molybdenum	18.7	+/-100	U	200	P	09/16/2025	14:47	LB137209
	Nickel	3.06	+/-20	U	40.0	P	09/16/2025	14:47	LB137209
	Silver	1.62	+/-5	U	10.0	P	09/16/2025	14:47	LB137209
	Zinc	16.7	+/-20	U	40.0	P	09/16/2025	14:47	LB137209
CCB04	Cadmium	0.50	+/-3	U	6.00	P	09/16/2025	15:36	LB137209
	Chromium	2.12	+/-5	U	10.0	P	09/16/2025	15:36	LB137209
	Copper	4.60	+/-10	U	20.0	P	09/16/2025	15:36	LB137209
	Lead	2.30	+/-6	U	12.0	P	09/16/2025	15:36	LB137209
	Molybdenum	18.7	+/-100	U	200	P	09/16/2025	15:36	LB137209
	Nickel	3.06	+/-20	U	40.0	P	09/16/2025	15:36	LB137209
	Silver	1.62	+/-5	U	10.0	P	09/16/2025	15:36	LB137209
	Zinc	16.7	+/-20	U	40.0	P	09/16/2025	15:36	LB137209
CCB05	Cadmium	0.50	+/-3	U	6.00	P	09/16/2025	16:31	LB137209
	Chromium	2.12	+/-5	U	10.0	P	09/16/2025	16:31	LB137209
	Copper	4.60	+/-10	U	20.0	P	09/16/2025	16:31	LB137209
	Lead	2.30	+/-6	U	12.0	P	09/16/2025	16:31	LB137209
	Molybdenum	18.7	+/-100	U	200	P	09/16/2025	16:31	LB137209
	Nickel	3.06	+/-20	U	40.0	P	09/16/2025	16:31	LB137209
	Silver	1.62	+/-5	U	10.0	P	09/16/2025	16:31	LB137209

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Elegant Jewelers Mfg. Co. Inc.

SDG No.: Q3063

Contract: ELEG01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB05	Zinc	16.7	+/-20	U	40.0	P	09/16/2025	16:31	LB137209
CCB06	Cadmium	0.50	+/-3	U	6.00	P	09/16/2025	16:53	LB137209
	Chromium	2.12	+/-5	U	10.0	P	09/16/2025	16:53	LB137209
	Copper	4.60	+/-10	U	20.0	P	09/16/2025	16:53	LB137209
	Lead	2.30	+/-6	U	12.0	P	09/16/2025	16:53	LB137209
	Molybdenum	18.7	+/-100	U	200	P	09/16/2025	16:53	LB137209
	Nickel	3.06	+/-20	U	40.0	P	09/16/2025	16:53	LB137209
	Silver	1.62	+/-5	U	10.0	P	09/16/2025	16:53	LB137209
	Zinc	16.7	+/-20	U	40.0	P	09/16/2025	16:53	LB137209

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Elegant Jewelers Mfg. Co. Inc.

SDG No.: Q3063

Contract: ELEG01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cadmium	0.50	+/-3	U	6.00	P	09/19/2025	12:25	LB137251
	Chromium	2.12	+/-5	U	10.0	P	09/19/2025	12:25	LB137251
	Copper	4.60	+/-10	U	20.0	P	09/19/2025	12:25	LB137251
	Lead	2.30	+/-6	U	12.0	P	09/19/2025	12:25	LB137251
	Molybdenum	18.7	+/-100	U	200	P	09/19/2025	12:25	LB137251
	Nickel	3.06	+/-20	U	40.0	P	09/19/2025	12:25	LB137251
	Silver	1.62	+/-5	U	10.0	P	09/19/2025	12:25	LB137251
	Zinc	16.7	+/-20	U	40.0	P	09/19/2025	12:25	LB137251

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Elegant Jewelers Mfg. Co. Inc.

SDG No.: Q3063

Contract: ELEG01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cadmium	0.50	+/-3	U	6.00	P	09/19/2025	13:13	LB137251
	Chromium	2.12	+/-5	U	10.0	P	09/19/2025	13:13	LB137251
	Copper	4.60	+/-10	U	20.0	P	09/19/2025	13:13	LB137251
	Lead	2.30	+/-6	U	12.0	P	09/19/2025	13:13	LB137251
	Molybdenum	18.7	+/-100	U	200	P	09/19/2025	13:13	LB137251
	Nickel	3.06	+/-20	U	40.0	P	09/19/2025	13:13	LB137251
	Silver	1.62	+/-5	U	10.0	P	09/19/2025	13:13	LB137251
	Zinc	16.7	+/-20	U	40.0	P	09/19/2025	13:13	LB137251
CCB02	Cadmium	0.50	+/-3	U	6.00	P	09/19/2025	14:40	LB137251
	Chromium	2.12	+/-5	U	10.0	P	09/19/2025	14:40	LB137251
	Copper	4.60	+/-10	U	20.0	P	09/19/2025	14:40	LB137251
	Lead	2.30	+/-6	U	12.0	P	09/19/2025	14:40	LB137251
	Molybdenum	18.7	+/-100	U	200	P	09/19/2025	14:40	LB137251
	Nickel	3.06	+/-20	U	40.0	P	09/19/2025	14:40	LB137251
	Silver	1.62	+/-5	U	10.0	P	09/19/2025	14:40	LB137251
	Zinc	16.7	+/-20	U	40.0	P	09/19/2025	14:40	LB137251
CCB03	Cadmium	0.50	+/-3	U	6.00	P	09/19/2025	15:50	LB137251
	Chromium	2.12	+/-5	U	10.0	P	09/19/2025	15:50	LB137251
	Copper	4.60	+/-10	U	20.0	P	09/19/2025	15:50	LB137251
	Lead	2.30	+/-6	U	12.0	P	09/19/2025	15:50	LB137251
	Molybdenum	18.7	+/-100	U	200	P	09/19/2025	15:50	LB137251
	Nickel	3.06	+/-20	U	40.0	P	09/19/2025	15:50	LB137251
	Silver	1.62	+/-5	U	10.0	P	09/19/2025	15:50	LB137251
	Zinc	16.7	+/-20	U	40.0	P	09/19/2025	15:50	LB137251
CCB04	Cadmium	0.50	+/-3	U	6.00	P	09/19/2025	17:12	LB137251
	Chromium	2.12	+/-5	U	10.0	P	09/19/2025	17:12	LB137251
	Copper	4.60	+/-10	U	20.0	P	09/19/2025	17:12	LB137251
	Lead	2.30	+/-6	U	12.0	P	09/19/2025	17:12	LB137251
	Molybdenum	18.7	+/-100	U	200	P	09/19/2025	17:12	LB137251
	Nickel	3.06	+/-20	U	40.0	P	09/19/2025	17:12	LB137251
	Silver	1.62	+/-5	U	10.0	P	09/19/2025	17:12	LB137251
	Zinc	16.7	+/-20	U	40.0	P	09/19/2025	17:12	LB137251