

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Elegant Jewelers Mfg. Co. Inc.</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3063</u>
Contract:	<u>ELEG01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3063-02</u>	Client ID:	<u>FALL-2025DUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3063-02DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	0.20	U	0.15	J	200.0		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Elegant Jewelers Mfg. Co. Inc.</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3063</u>
Contract:	<u>ELEG01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3063-02MS</u>	Client ID:	<u>FALL-2025MSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3063-02MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	4.02		4.08		1		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Elegant Jewelers Mfg. Co. Inc. **Level:** LOW **SDG No.:** Q3063
Contract: ELEG01 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q3098-04 **Client ID:** EFF-WWDUP
Percent Solids for Sample: NA **Duplicate ID** Q3098-04DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cadmium	ug/L	20	3.00	U	3.00	U			P
Chromium	ug/L	20	1.53	J	1.06	J	36		P
Copper	ug/L	20	98.8		96.2		3		P
Lead	ug/L	20	1.31	J	1.25	J	5		P
Molybdenum	ug/L	20	100	U	100	U			P
Nickel	ug/L	20	2.09	J	2.23	J	6		P
Silver	ug/L	20	5.00	U	5.00	U			P
Zinc	ug/L	20	128		130		2		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Elegant Jewelers Mfg. Co. Inc. **Level:** LOW **SDG No.:** Q3063
Contract: ELEG01 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q3098-05 **Client ID:** EFF-WWMSD
Percent Solids for Sample: NA **Duplicate ID** Q3098-06 **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cadmium	ug/L	20	106		100		6		P
Chromium	ug/L	20	220		208		6		P
Copper	ug/L	20	260		244		6		P
Lead	ug/L	20	508		480		6		P
Molybdenum	ug/L	20	221		212		4		P
Nickel	ug/L	20	270		255		6		P
Silver	ug/L	20	21.2		21.5		1		P
Zinc	ug/L	20	249		232		7		P