



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Ambric Technology Corporation

Project Name: NEWPCP Sludge Screen 13923-GE-25

Project # N/A

Order ID # Q3068

Test Name: Metals Group4

A. Number of Samples and Date of Receipt:

1 Solid sample was received on 09/10/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Metals Group4. This data package contains results for Metals Group4.

C. Analytical Techniques:

The analysis of Metals Group4 was based on method 6010D and digestion based on method 3050 (soils).

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all compounds.

The Duplicate (MH-384DUP) analysis met criteria for all compounds except for Calcium, Copper, Manganese and Zinc due to Sample Matrix interference.

The Matrix Spike analysis met criteria for all compounds.

The Matrix Spike Duplicate analysis met criteria for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

E. Additional Comments:

The Date and time of sampling was not listed in the COC.

The temperature of the samples at the time of receipt was 27.3°C.

The Fax and Hardcopy is not matching. At the time of second review lab noticed the fax data were not calculated with Percent solids, therefore the data were reprocessed with Percent solids and reported in Hardcopy. Hardcopy is reported corrected.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____