



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Ambric Technology Corporation

Project Name: JF Huebner

Project # N/A

Order ID # Q3068

Test Name: Anions Group2,pH,Phosphorus, Total,TOC

A. Number of Samples and Date of Receipt:

1 Solid sample was received on 09/10/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Anions Group2,pH,Phosphorus, Total,TOC. This data package contains results for Anions Group2,pH,Phosphorus, Total,TOC.

C. Analytical Techniques:

The analysis of pH was based on method 9045D, The analysis of Anions Group2 was based on method 9056A and The analysis of TOC was based on method 9060A, The analysis of Phosphorus, Total was based on method 365.3

D. QA/ QC Samples:

The Holding Times were met for all samples except for AMBRIC-TECH-9-9 of pH as sample was receive out of holding time.

Sample AMBRIC-TECH-9-9 was diluted due to high concentrations for Phosphorus, Total.

The Blank Spike met requirements for all compounds.

The Duplicate analysis met criteria for all compounds.

The Matrix Spike (AMBRIC-TECH-9-9MS) analysis met criteria for all compounds except for Chloride and Phosphorus, Total due to sample matrix interference.

The Matrix Spike Duplicate (AMBRIC-TECH-9-9MSD) analysis met criteria for all compounds except for Chloride and Phosphorus, Total due to sample matrix interference.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

E. Additional Comments:

This package has been revised due to project name updated.

Lab has used least representable sample weight for the sample Q3068-01 for TOC analysis. Therefore Lab has reported the TOC result with "OR" qualifier.

The Date and time of sampling was not listed in the COC.

The temperature of the samples at the time of receipt was 27.3°C.



I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____