

**ALLIANCE** 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

**GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY**

ORDER ID: Q3068

MATRIX: Solid

METHOD: 365.3,9045D,9056A,9060A

	NA	NO	YES
1. Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
2. Matrix Spike Duplicate Recoveries Met Criteria		✓	
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
The Blank Spike met requirements for all compounds. The Matrix Spike (AMBRIC-TECH-9-9MS) analysis met criteria for all compounds except for Chloride and Phosphorus, Total due to sample matrix interference. The Matrix Spike Duplicate (AMBRIC-TECH-9-9MSD) analysis met criteria for all compounds except for Chloride and Phosphorus, Total due to sample matrix interference.			
3. Sample Duplicate Analysis Met QC Criteria			✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
4. Digestion Holding Time Met		✓	
If not met, list number of days exceeded for each sample:			
The Holding Times were met for all samples except for AMBRIC-TECH-9-9 of pH as sample was receive out of holding time.			

**ADDITIONAL COMMENTS:**

This package has been revise due to project name updated.

Lab has used least representable sample weight for the sample Q3068-01 for TOC analysis. Therefore Lab has reported the TOC result with "OR" qualifier.

The Date and time of sampling was not listed in the COC.

The temperature of the samples at the time of receipt was 27.3°C.

\_\_\_\_\_  
QA REVIEW

\_\_\_\_\_  
Date