

**ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092**

NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

**METALS CONFORMANCE/NON-CONFORMANCE SUMMARY**

ORDER ID: Q3068

MATRIX: Solid

METHOD: 6010D

	NA	NO	YES
1. Calibration Summary met criteria.			✓
2. ICP Interference Check Sample Results Summary Submitted.			✓
3. Serial Dilution Summary (if applicable) Submitted.			✓
4. Laboratory Control Sample Summary (if applicable) Submitted.			✓
5. Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
6. Matrix Spike/Matrix Spike Duplicate Recoveries Met Criteria			✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
7. Sample Duplicate Analysis Met QC Criteria		✓	
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
The Duplicate (MH-384DUP) analysis met criteria for all compounds except for Calcium, Copper, Manganese and Zinc due to Sample Matrix interference.			
8. Digestion Holding Time Met			✓
If not met, list number of days exceeded for each sample:			
9. Analysis Holding Time Met			✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.			

**ADDITIONAL COMMENTS:**

This package has been revise due to project name updated.

The Date and time of sampling was not listed in the COC.

The temperature of the samples at the time of receipt was 27.3°C.

The Fax and Hardcopy is not matching. At the time of second review lab noticed the fax data were not calculated with Percent solids, therefore the data were reprocessed with Percent solids and reported in Hardcopy. Hardcopy is reported corrected.

---

QA REVIEW

---

Date