

Instrument ID: IC-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137133

Review By	rubina	Review On	9/11/2025 9:34:39 AM
Supervise By	Iwona	Supervise On	9/11/2025 12:56:45 PM
SubDirectory	LB137133	Test	Anions
STD. NAME	STD REF.#		
ICAL Standard	WP114677,WP114678,WP114679,WP114680,WP114681,WP114682,WP114683		
ICV Standard	WP114684		
CCV Standard	WP114699		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP114700		
Chk Standard	WP114685,WP114686		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	STD1	STD1	CAL1	09/09/25 10:21	All standards, samples, and	RM/IZ	OK
2	STD2	STD2	CAL2	09/09/25 10:42	QC are filtered through	RM/IZ	OK
3	STD3	STD3	CAL3	09/09/25 11:25	0.45um, filter lot W3160	RM/IZ	OK
4	STD4	STD4	CAL4	09/09/25 11:47		RM/IZ	OK
5	STD5	STD5	CAL5	09/09/25 13:36		RM/IZ	OK
6	STD6	STD6	CAL6	09/09/25 13:58		RM/IZ	OK
7	STD7	STD7	CAL7	09/09/25 14:41		RM/IZ	OK
8	ICV1	ICV1	ICV	09/09/25 15:02		RM/IZ	OK
9	ICB1	ICB1	ICB	09/09/25 15:45		RM/IZ	OK
10	CCV1	CCV1	CCV	09/10/25 11:08		RM/IZ	OK
11	CCB1	CCB1	CCB	09/10/25 11:29		RM/IZ	OK
12	LB137133BLS	LB137133BLS	MB	09/10/25 11:51		RM/IZ	OK
13	LB137133BSS	LB137133BSS	LCS	09/10/25 12:12		RM/IZ	OK
14	Q3068-01	AMBRIC-TECH-9-9	SAM	09/10/25 12:34		RM/IZ	OK
15	Q3068-01MS	AMBRIC-TECH-9-9MS	MS	09/10/25 12:55	5ml W3092	RM/IZ	OK
16	Q3068-01MSD	AMBRIC-TECH-9-9MSD	MSD	09/10/25 13:17	5ml W3092	RM/IZ	OK
17	CCV2	CCV2	CCV	09/10/25 13:38		RM/IZ	OK
18	CCB2	CCB2	CCB	09/10/25 14:00		RM/IZ	OK

Instrument ID: TOC

Daily Analysis Runlog For Sequence/QC Batch ID # LB137140

Review By	rubina	Review On	9/11/2025 9:38:24 AM
Supervise By	Iwona	Supervise On	9/11/2025 1:00:12 PM
SubDirectory	LB137140	Test	TOC
STD. NAME	STD REF.#		
ICAL Standard	WP114036,WP114037,WP114038,WP114039,WP114040		
ICV Standard	WP114041		
CCV Standard	WP114722		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP114723		
Chk Standard	WP112446,WP112447		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	BLANK	BLANK	CAL1	07/22/25 10:12		RM IZ	OK
2	250mg/l	250mg/l	CAL2	07/22/25 10:21		RM IZ	OK
3	500mg/l	500mg/l	CAL3	07/22/25 10:29		RM IZ	OK
4	1000mg/l	1000mg/l	CAL4	07/22/25 10:39		RM IZ	OK
5	2000mg/l	2000mg/l	CAL5	07/22/25 10:50		RM IZ	OK
6	ICV1	ICV1	ICV	07/22/25 11:03		RM IZ	OK
7	ICB1	ICB1	ICB	07/22/25 11:19		RM IZ	OK
8	CCV1	CCV1	CCV	09/10/25 12:40		RM IZ	OK
9	CCB1	CCB1	CCB	09/10/25 12:54		RM IZ	OK
10	LB137140BLS	LB137140BLS	MB	09/10/25 13:12		RM IZ	OK
11	LB137140BSS	LB137140BSS	LCS	09/10/25 13:30		RM IZ	OK
12	Q3068-01	AMBRIC-TECH-9-9	SAM	09/10/25 13:53		RM IZ	OK
13	Q3068-01MS	AMBRIC-TECH-9-9MS	MS	09/10/25 14:07	sample + 40ul of WP114722	RM IZ	OK
14	Q3068-01MSD	AMBRIC-TECH-9-9MSD	MSD	09/10/25 14:39	sample + 40ul of WP114722	RM IZ	OK
15	CCV2	CCV2	CCV	09/10/25 15:00		RM IZ	OK
16	CCB2	CCB2	CCB	09/10/25 15:14		RM IZ	OK

Instrument ID: WC PH METER-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137148

Review By	JIGNESH	Review On	9/11/2025 10:17:03 AM
Supervise By	Iwona	Supervise On	9/11/2025 11:13:19 AM
SubDirectory	LB137148	Test	pH
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	W3178,W3093,W3191,W3217,W3161,W3200		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	CAL1	CAL1	CAL	09/11/25 08:30		Jignesh	OK
2	CAL2	CAL2	CAL	09/11/25 08:31		Jignesh	OK
3	CAL3	CAL3	CAL	09/11/25 08:33		Jignesh	OK
4	ICV	ICV	ICV	09/11/25 08:35		Jignesh	OK
5	CCV1	CCV1	CCV	09/11/25 08:38		Jignesh	OK
6	Q3068-01	AMBRIC-TECH-9-9	SAM	09/11/25 08:45		Jignesh	OK
7	Q3068-01DUP	AMBRIC-TECH-9-9DUP	DUP	09/11/25 08:47		Jignesh	OK
8	CCV2	CCV2	CCV	09/11/25 08:50		Jignesh	OK

Instrument ID: SPECTROPHOTOMETER-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137156

Review By	Iwona	Review On	9/11/2025 2:33:06 PM
Supervise By	JIGNESH	Supervise On	9/11/2025 2:36:59 PM
SubDirectory	LB137156	Test	Phosphorus, Total
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	WP114614,WP114613,WP114612,WP114611,WP114610,WP114609,WP114615,WP112831,WP114732,WP113378,V		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	CAL1	CAL1	CAL	09/11/25 13:15		Iwona	OK
2	CAL2	CAL2	CAL	09/11/25 13:15		Iwona	OK
3	CAL3	CAL3	CAL	09/11/25 13:16		Iwona	OK
4	CAL4	CAL4	CAL	09/11/25 13:16		Iwona	OK
5	CAL5	CAL5	CAL	09/11/25 13:17		Iwona	OK
6	CAL6	CAL6	CAL	09/11/25 13:17		Iwona	OK
7	ICV	ICV	ICV	09/11/25 13:18		Iwona	OK
8	ICB	ICB	ICB	09/11/25 13:18		Iwona	OK
9	CCV1	CCV1	CCV	09/11/25 13:19		Iwona	OK
10	CCB1	CCB1	CCB	09/11/25 13:19		Iwona	OK
11	RL Check	RL Check	RL	09/11/25 13:20		Iwona	OK
12	PB169664BL	PB169664BL	MB	09/11/25 13:20		Iwona	OK
13	PB169664BS	PB169664BS	LCS	09/11/25 13:21		Iwona	OK
14	Q2893-03	MDL-SOIL-03-QT3-20	SAM	09/11/25 13:21		Iwona	OK
15	Q3068-01	AMBRIC-TECH-9-9	SAM	09/11/25 13:22		Iwona	OK
16	Q3068-01DUP	AMBRIC-TECH-9-9DUP	DUP	09/11/25 13:22		Iwona	OK
17	Q3068-01MS	AMBRIC-TECH-9-9MS	MS	09/11/25 13:23		Iwona	OK
18	Q3068-01MSD	AMBRIC-TECH-9-9MSD	MSD	09/11/25 13:23		Iwona	OK

Instrument ID: SPECTROPHOTOMETER-1

Daily Analysis Runlog For Sequence/QCBatch ID # LB137156

Review By	Iwona	Review On	9/11/2025 2:33:06 PM
Supervise By	JIGNESH	Supervise On	9/11/2025 2:36:59 PM
SubDirectory	LB137156	Test	Phosphorus, Total
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	WP114614,WP114613,WP114612,WP114611,WP114610,WP114609,WP114615,WP112831,WP114732,WP113378,V		

19	Q3068-01DL	AMBRIC-TECH-9-9DL	SAM	09/11/25 13:24		Iwona	OK
20	Q3068-01DUP	AMBRIC-TECH-9-9DL	DUP	09/11/25 13:24		Iwona	OK
21	CCV2	CCV2	CCV	09/11/25 13:25		Iwona	OK
22	CCB2	CCB2	CCB	09/11/25 13:25		Iwona	OK
23	Q3068-01MS	AMBRIC-TECH-9-9MS	MS	09/11/25 13:26		Iwona	OK
24	Q3068-01MSD	AMBRIC-TECH-9-9MS	MSD	09/11/25 13:26		Iwona	OK
25	CCV3	CCV3	CCV	09/11/25 13:27		Iwona	OK
26	CCB3	CCB3	CCB	09/11/25 13:27		Iwona	OK