

Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB137148

Slope : 98.4

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/11/2025	08:30
2	CAL2	1	Water	NA	NA	20.2	7.00	09/11/2025	08:31
3	CAL3	1	Water	NA	NA	20.2	10.02	09/11/2025	08:33
4	ICV	1	Water	NA	NA	20.3	7.02	09/11/2025	08:35
5	CCV1	1	Water	NA	NA	20.2	2.01	09/11/2025	08:38
6	Q3068-01	1	Solid	20.02	20	20.2	7.80	09/11/2025	08:45
7	Q3068-01DUP	1	Solid	20.03	20	20.3	7.82	09/11/2025	08:47
8	CCV2	1	Water	NA	NA	20.2	12.02	09/11/2025	08:50

WORKLIST(Hardcopy Internal Chain)

137148

WorkList Name : ph s q3068

WorkList ID : 191796

Department : Wet-Chemistry

Date : 09-11-2025 08:14:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3068-01	B AMBRIC-TECH-9-9	Solid	pH	Cool 4 deg C	AMBR01	D31	09/09/2025	9045D

Date/Time 09-11-25 08:25

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 09-11-25

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

13:00

[Signature]

[Signature]