

Initial and Continuing Calibration Blank Summary

Client: Ambric Technology Corporation

SDG No.: Q3068

Project: NEWPCP Sludge Screen 13923-GE-25

RunNo.: LB137133

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	09/09/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	09/09/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	09/09/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	09/09/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	09/09/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	09/09/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	09/09/2025
Sample ID: CCB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	09/10/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	09/10/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	09/10/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	09/10/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	09/10/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	09/10/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	09/10/2025
Sample ID: CCB2							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	09/10/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	09/10/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	09/10/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	09/10/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	09/10/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	09/10/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	09/10/2025

Initial and Continuing Calibration Blank Summary

Client: Ambric Technology Corporation

SDG No.: Q3068

Project: NEWPCP Sludge Screen 13923-GE-25

RunNo.: LB137140

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 TOC	mg/L	< 125.0000	125.0000	U	22.3	250	07/22/2025
Sample ID: CCB1 TOC	mg/L	< 125.0000	125.0000	U	22.3	250	09/10/2025
Sample ID: CCB2 TOC	mg/L	< 125.0000	125.0000	U	22.3	250	09/10/2025

Initial and Continuing Calibration Blank Summary

Client: Ambric Technology Corporation

SDG No.: Q3068

Project: NEWPCP Sludge Screen 13923-GE-25

RunNo.: LB137156

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Phosphorus, Total	mg/L	0.008	0.0250	J	0.0045	0.05	09/11/2025
Sample ID: CCB1 Phosphorus, Total	mg/L	0.007	0.0250	J	0.0045	0.05	09/11/2025
Sample ID: CCB2 Phosphorus, Total	mg/L	0.005	0.0250	J	0.0045	0.05	09/11/2025
Sample ID: CCB3 Phosphorus, Total	mg/L	0.008	0.0250	J	0.0045	0.05	09/11/2025