

Initial and Continuing Calibration Verification

Client: Ambric Technology Corporation

SDG No.: Q3068

Project: JF Huebner

RunNo.: LB137133

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	10	10	100	90-110	09/09/2025
Chloride	mg/L	3	3	100	90-110	09/09/2025
Fluoride	mg/L	2	2	100	90-110	09/09/2025
Nitrite	mg/L	3	3	100	90-110	09/09/2025
Nitrate	mg/L	2.5	2.5	100	90-110	09/09/2025
Sulfate	mg/L	14.8	15	99	90-110	09/09/2025
Orthophosphate as P	mg/L	5	5	100	90-110	09/09/2025
Sample ID: CCV1						
Bromide	mg/L	10	10	100	90-110	09/10/2025
Chloride	mg/L	3	3	100	90-110	09/10/2025
Fluoride	mg/L	2	2	100	90-110	09/10/2025
Nitrite	mg/L	3	3	100	90-110	09/10/2025
Nitrate	mg/L	2.4	2.5	96	90-110	09/10/2025
Sulfate	mg/L	14.9	15	99	90-110	09/10/2025
Orthophosphate as P	mg/L	5	5	100	90-110	09/10/2025
Sample ID: CCV2						
Bromide	mg/L	10	10	100	90-110	09/10/2025
Chloride	mg/L	3	3	100	90-110	09/10/2025
Fluoride	mg/L	2	2	100	90-110	09/10/2025
Nitrite	mg/L	3	3	100	90-110	09/10/2025
Nitrate	mg/L	2.4	2.5	96	90-110	09/10/2025
Sulfate	mg/L	14.8	15	99	90-110	09/10/2025
Orthophosphate as P	mg/L	5.1	5	102	90-110	09/10/2025

Initial and Continuing Calibration Verification

Client: Ambric Technology Corporation

SDG No.: Q3068

Project: JF Huebner

RunNo.: LB137140

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 TOC	mg/L	968	1000	97	90-110	07/22/2025
Sample ID: CCV1 TOC	mg/L	1050	1000	105	90-110	09/10/2025
Sample ID: CCV2 TOC	mg/L	1080	1000	108	90-110	09/10/2025

Initial and Continuing Calibration Verification

Client: Ambric Technology Corporation

SDG No.: Q3068

Project: JF Huebner

RunNo.: LB137148

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.02	7	100	90-110	09/11/2025
Sample ID: CCV1 pH	pH	2.01	2.00	101	90-110	09/11/2025
Sample ID: CCV2 pH	pH	12.02	12.00	100	90-110	09/11/2025

Initial and Continuing Calibration Verification

Client: Ambric Technology Corporation

SDG No.: Q3068

Project: JF Huebner

RunNo.: LB137156

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.497	0.50	99	90-110	09/11/2025
Sample ID: CCV1 Phosphorus, Total	mg/L	0.491	0.50	98	90-110	09/11/2025
Sample ID: CCV2 Phosphorus, Total	mg/L	0.501	0.50	100	90-110	09/11/2025
Sample ID: CCV3 Phosphorus, Total	mg/L	0.494	0.50	99	90-110	09/11/2025