

Duplicate Sample Summary

Client: Ambric Technology Corporation	SDG No.: Q3068
Project: JF Huebner	Sample ID: Q3068-01
Client ID: AMBRIC-TECH-9-9DUP	Percent Solids for Spike Sample: 100

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
pH	pH	+/-20	7.80		7.82		1	0.26		09/11/2025

Duplicate Sample Summary

Client:	Ambric Technology Corporation	SDG No.:	Q3068
Project:	JF Huebner	Sample ID:	Q3068-01DL
Client ID:	AMBRIC-TECH-9-9DUP	Percent Solids for Spike Sample:	89

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Phosphorus, Total	mg/Kg	+/-20	303	D	297	D	10	2		09/11/2025

Duplicate Sample Summary

Client: Ambric Technology Corporation	SDG No.: Q3068
Project: JF Huebner	Sample ID: Q3068-01
Client ID: AMBRIC-TECH-9-9MSD	Percent Solids for Spike Sample: 89

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Bromide	mg/Kg	+/-15	226		227		1	0		09/10/2025
Chloride	mg/Kg	+/-15	251	OR	252	OR	1	0		09/10/2025
Nitrate	mg/Kg	+/-15	54.9		55.3		1	1		09/10/2025
Nitrite	mg/Kg	+/-15	66.7		67.3		1	1		09/10/2025
Sulfate	mg/Kg	+/-15	572		577		1	1		09/10/2025
Orthophosphate as P	mg/Kg	+/-15	94.4		108		1	13		09/10/2025
Fluoride	mg/Kg	+/-15	45.7		46.7		1	2		09/10/2025
TOC	mg/Kg	+/-20	16000	OR	16100	OR	1	1		09/10/2025
Phosphorus, Total	mg/Kg	+/-20	326	D	325	D	10	0.31		09/11/2025