

Laboratory Control Sample Summary

Client: Ambric Technology Corporation

SDG No.: Q3068

Project: JF Huebner

Run No.: LB137133

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137133BSS							
Bromide	mg/Kg	200	200		100	1	90-110	09/10/2025
Chloride	mg/Kg	60	59.0		98	1	90-110	09/10/2025
Fluoride	mg/Kg	40	40.3		101	1	90-110	09/10/2025
Nitrite	mg/Kg	60	59.3		99	1	90-110	09/10/2025
Nitrate	mg/Kg	50	48.5		97	1	90-110	09/10/2025
Sulfate	mg/Kg	300	297		99	1	90-110	09/10/2025
Orthophosphate as P	mg/Kg	100	100		100	1	90-110	09/10/2025

Laboratory Control Sample Summary

Client: Ambric Technology Corporation

SDG No.: Q3068

Project: JF Huebner

Run No.: LB137140

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137140BSS							
TOC	mg/Kg	1000	1040		104	1	90-110	09/10/2025

Laboratory Control Sample Summary

Client: Ambric Technology Corporation

SDG No.: Q3068

Project: JF Huebner

Run No.: LB137156

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169664BS							
Phosphorus, Total	mg/Kg	25.0	25.0		100	1	90-110	09/11/2025