



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Ambric Technology Corporation

SDG No.: Q3068

Contract: AMBR01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Calcium	2080	2000	104	65 - 135	P	09/10/2025	11:48	LB137144
	Copper	21.7	20.0	108	65 - 135	P	09/10/2025	11:48	LB137144
	Iron	109	100	109	65 - 135	P	09/10/2025	11:48	LB137144
	Magnesium	2100	2000	105	65 - 135	P	09/10/2025	11:48	LB137144
	Manganese	21.5	20.0	108	65 - 135	P	09/10/2025	11:48	LB137144
	Potassium	1970	2000	99	65 - 135	P	09/10/2025	11:48	LB137144
	Zinc	43.0	40.0	108	65 - 135	P	09/10/2025	11:48	LB137144