

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client:	<u>Ambric Technology Corporation</u>	SDG No.:	<u>Q3068</u>
Contract:	<u>AMBR01</u>	Lab Code:	<u>ACE</u>
ICS Source:	<u>EPA</u>	Instrument ID:	<u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Calcium	265000	245000	108	208000	282000	09/10/2025	11:52	LB137144
	Copper	2.32	2.0	116	-18	22	09/10/2025	11:52	LB137144
	Iron	111000	101000	110	85600	116500	09/10/2025	11:52	LB137144
	Magnesium	281000	255000	110	216000	294000	09/10/2025	11:52	LB137144
	Manganese	10.8	7.0	155	-13	27	09/10/2025	11:52	LB137144
	Potassium	200			0	0	09/10/2025	11:52	LB137144
	Zinc	4.41			-40	40	09/10/2025	11:52	LB137144
ICSAB01	Calcium	236000	235000	100	199000	271000	09/10/2025	11:56	LB137144
	Copper	471	511	92	434	588	09/10/2025	11:56	LB137144
	Iron	97500	99300	98	84400	114500	09/10/2025	11:56	LB137144
	Magnesium	251000	248000	101	210000	286000	09/10/2025	11:56	LB137144
	Manganese	486	507	96	430	584	09/10/2025	11:56	LB137144
	Potassium	-95.0			0	0	09/10/2025	11:56	LB137144
	Zinc	1030	952	108	809	1095	09/10/2025	11:56	LB137144