

## Initial and Continuing Calibration Verification

**Client:** Major Products

**SDG No.:** Q3070

**Project:** Floor Drain Water 2025

**RunNo.:** LB137149

| Analyte                      | Units | Result | True Value | % Recovery | Acceptance Window (%R) | Analysis Date |
|------------------------------|-------|--------|------------|------------|------------------------|---------------|
| Sample ID: <b>ICV</b><br>pH  | pH    | 7.02   | 7          | 100        | 90-110                 | 09/11/2025    |
| Sample ID: <b>CCV1</b><br>pH | pH    | 2.02   | 2.00       | 101        | 90-110                 | 09/11/2025    |
| Sample ID: <b>CCV2</b><br>pH | pH    | 12.02  | 12.00      | 100        | 90-110                 | 09/11/2025    |