

Duplicate Sample Summary

Client: Major Products Project: Floor Drain Water 2025 Client ID: LB137183BSD	SDG No.: Q3070 Sample ID: LB137183BS Percent Solids for Spike Sample: 0
--	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TPH	mg/L	+/-18	17.1		16.9		1	1.18		09/15/2025

Duplicate Sample Summary

Client: Major Products Project: Floor Drain Water 2025 Client ID: COMPDUP	SDG No.: Q3070 Sample ID: Q3067-02 Percent Solids for Spike Sample: 0
--	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	681		683		1	0.24		09/10/2025
TSS	mg/L	+/-5	241		238		1	1.25		09/15/2025

Duplicate Sample Summary

Client:	Major Products	SDG No.:	Q3070
Project:	Floor Drain Water 2025	Sample ID:	Q3070-01
Client ID:	DRAIN-WATER-TANK-1MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	43.9		44.0		1	0.23		09/15/2025

Duplicate Sample Summary

Client: Major Products Project: Floor Drain Water 2025 Client ID: MOO-25-0140DUP	SDG No.: Q3070 Sample ID: Q3074-02 Percent Solids for Spike Sample: 0
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
pH	pH	+/-20	6.30		6.32		1	0.32		09/11/2025

Duplicate Sample Summary

Client:	Major Products	SDG No.:	Q3070
Project:	Floor Drain Water 2025	Sample ID:	Q3081-01
Client ID:	DRAIN-WATER-TANK-1MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	22.7		23.0		1	1.31		09/15/2025

Duplicate Sample Summary

Client: Major Products Project: Floor Drain Water 2025 Client ID: MH-9-12-25MSD	SDG No.: Q3070 Sample ID: Q3093-01 Percent Solids for Spike Sample: 0
--	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	149		149		1	0.07		09/15/2025

Duplicate Sample Summary

Client: Major Products	SDG No.: Q3070
Project: Floor Drain Water 2025	Sample ID: Q3096-01
Client ID: DRAIN-WATER-TANK-1MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	25.7		25.8		1	0.39		09/15/2025