

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>GARD04</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3078</u>
Instrument ID:	<u>MSVOA_N</u>	Calibration Date/Time:	<u>09/12/2025</u> <u>10:04</u>
Lab File ID:	<u>VN087811.D</u>	Init. Calib. Date(s):	<u>09/04/2025</u> <u>09/04/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>16:23</u> <u>17:46</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Acrolein	0.497	0.474		-4.63	
Acrylonitrile	1.202	0.967		-19.55	
1,2-Dichloroethane-d4	2.563	2.485	0.01	-3.04	
Toluene-d8	1.395	1.380	0.01	-1.08	
4-Bromofluorobenzene	0.482	0.491	0.2	1.87	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.