

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GARD04
 Lab Code: ACE SDG NO.: Q3078
 Lab File ID: VN087811.D Date Analyzed: 09/12/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:04
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	60784	7.80	303424	9.08	286682	11.85
UPPER LIMIT	121568	8.3	606848	9.582	573364	12.347
LOWER LIMIT	30392	7.3	151712	8.582	143341	11.347
EPA SAMPLE NO.						
250910074-01-VOA	55446	7.79	279587	9.08	261896	11.85
250910064-04-TRIP-BLANK	47676	7.80	237316	9.08	222106	11.85
VN0912WBL01	57083	7.80	279905	9.08	254454	11.84
VN0912WBS01	57185	7.79	292500	9.08	270542	11.84
VN0912WBSD01	60654	7.80	293305	9.08	272699	11.85

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.