

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN082125

Review By	John Carlone	Review On	8/22/2025 9:17:52 AM
Supervise By	Maresh Dadoda	Supervise On	8/25/2025 8:11:38 AM
SubDirectory	VN082125	HP Acquire Method	HP Processing Method 82N082125W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135214 VP135215,VP135216,VP135217,VP135218,VP135219,VP135220 VP135221

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087614.D	21 Aug 2025 09:30		JC\MD	Ok
2	VSTDCCC020	VSTDCCC020	VN087615.D	21 Aug 2025 10:04	Need ICAL	JC\MD	Not Ok
3	VN0821WBS01	VN0821WBS01	VN087616.D	21 Aug 2025 10:37		JC\MD	Not Ok
4	VN0821WBSD01	VN0821WBSD01	VN087617.D	21 Aug 2025 11:11		JC\MD	Not Ok
5	VN0821WBL01	VN0821WBL01	VN087618.D	21 Aug 2025 11:36		JC\MD	Not Ok
6	VSTDICC001	VSTDICC001	VN087619.D	21 Aug 2025 12:09	LR-10,20	JC\MD	Ok,M
7	VSTDICC005	VSTDICC005	VN087620.D	21 Aug 2025 13:08		JC\MD	Ok,M
8	VSTDICC020	VSTDICC020	VN087621.D	21 Aug 2025 13:41	method not used for #N-amyle Acetate	JC\MD	Ok,M
9	VSTDICCC050	VSTDICCC050	VN087622.D	21 Aug 2025 14:02		JC\MD	Ok,M
10	VSTDICC100	VSTDICC100	VN087623.D	21 Aug 2025 14:23		JC\MD	Ok,M
11	VSTDICC150	VSTDICC150	VN087624.D	21 Aug 2025 14:44		JC\MD	Ok,M
12	IBLK	IBLK	VN087625.D	21 Aug 2025 15:05		JC\MD	Ok
13	VSTDICV050	ICVVN082125	VN087626.D	21 Aug 2025 15:47	Icv Failed for Com.# 64 for DOD	JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN091125

Review By	Mahesh Dadoda	Review On	9/12/2025 1:48:33 PM
Supervise By	Semsettin Yesilyurt	Supervise On	9/12/2025 1:52:34 PM
SubDirectory	VN091125	HP Acquire Method	MSVOA_N
		HP Processing Method	VN082125W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135417
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135415,VP135416 VP134956

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087791.D	11 Sep 2025 09:53		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN087792.D	11 Sep 2025 12:49		JC\MD	Ok,M
3	VN0911MBL01	VN0911MBL01	VN087793.D	11 Sep 2025 13:25		JC\MD	Ok
4	VN0911WBL01	VN0911WBL01	VN087794.D	11 Sep 2025 13:45		JC\MD	Ok
5	VN0911WBS01	VN0911WBS01	VN087795.D	11 Sep 2025 14:21	BS Failed High For Com.#40,65,68	JC\MD	Ok,M
6	VN0911WBSD01	VN0911WBSD01	VN087796.D	11 Sep 2025 14:42	BSD Failed High For Com.#39,40,65,67,68,69	JC\MD	Ok,M
7	Q3078-01	250910074-01-VOA	VN087797.D	11 Sep 2025 15:03	Need 10X;BS,BSD Failed High,vial A pH#8	JC\MD	Dilution
8	VIBLK	VIBLK	VN087798.D	11 Sep 2025 15:25		JC\MD	Ok
9	Q3078-02	250910064-04-TRIP-BL	VN087799.D	11 Sep 2025 15:45	TB,vial A pH#7	JC\MD	Ok
10	PB169644TB	PB169644TB	VN087800.D	11 Sep 2025 16:06	vial A pH# 5	JC\MD	Ok,M
11	Q3054-04	MH-385	VN087801.D	11 Sep 2025 16:27	vial A pH# 5	JC\MD	Ok,M
12	Q3054-08	MH-384	VN087802.D	11 Sep 2025 16:48	vial B pH# 5	JC\MD	Ok
13	Q3072-02	30-GARFIELD-PL-COM	VN087803.D	11 Sep 2025 17:09	vial A pH# 5	JC\MD	Ok,M
14	Q3083-01	MW4S	VN087804.D	11 Sep 2025 17:30	BS failed High;BSD failed High,vial A pH#<2	JC\MD	ReRun
15	Q3083-02	MW4I	VN087805.D	11 Sep 2025 17:51	vial A pH#<3	JC\MD	Ok
16	Q3083-03	MW3S	VN087806.D	11 Sep 2025 18:11	vial A pH# <2	JC\MD	Ok
17	Q3078-01DL	250910074-01-VOADL	VN087807.D	11 Sep 2025 18:32	BS,BSD Failed High,vial C pH# 8,NO SAMPLE TO RERUN	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN091125

Review By	Mahesh Dadoda	Review On	9/12/2025 1:48:33 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	9/12/2025 1:52:34 PM		
SubDirectory	VN091125	HP Acquire Method	MSVOA_N	HP Processing Method	VN082125W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135417
CCC	VP135415,VP135416
Internal Standard/PEM	VP134956
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	VN0911WBS02	VN0911WBS02	VN087808.D	11 Sep 2025 18:53	Surrogate Fail	JC\MD	Not Ok
19	VSTDCCC050	VSTDCCC050EC	VN087809.D	11 Sep 2025 19:14		JC\MD	Ok,M

M : Manual Integration