

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GARD04
 Lab Code: ACE SDG NO.: Q3078
 Lab File ID: VN087792.D Date Analyzed: 09/11/2025
 Instrument ID: MSVOA_N Time Analyzed: 12:49
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	262340	13.77				
UPPER LIMIT	524680	14.27				
LOWER LIMIT	131170	13.27				
EPA SAMPLE NO.						
250910074-01-VOA	215998	13.77				
250910074-01-VOADL	182604	13.77				
250910064-04-TRIP-BLANK	178191	13.77				
VN0911WBL01	224924	13.77				
VN0911WBS01	243902	13.77				
VN0911WBSD01	252435	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.