

Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB137185

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/15/2025	08:10
2	CAL2	1	Water	NA	NA	20.1	7.00	09/15/2025	08:11
3	CAL3	1	Water	NA	NA	20.2	10.02	09/15/2025	08:15
4	ICV	1	Water	NA	NA	20.3	7.00	09/15/2025	08:20
5	CCV1	1	Water	NA	NA	20.2	2.02	09/15/2025	08:40
6	Q3081-01	1	Water	NA	NA	20.1	6.85	09/15/2025	08:45
7	Q3081-01DUP	1	Water	NA	NA	20.2	6.87	09/15/2025	08:47
8	Q3088-02	1	Water	NA	NA	22.6	6.50	09/15/2025	08:50
9	Q3088-02DUP	1	Water	NA	NA	22.8	6.52	09/15/2025	08:52
10	Q3090-01	1	Water	NA	NA	23.6	8.40	09/15/2025	09:00
11	Q3090-01DUP	1	Water	NA	NA	23.7	8.42	09/15/2025	09:03
12	Q3093-01	1	Water	NA	NA	24.4	5.01	09/15/2025	09:15
13	Q3093-01DUP	1	Water	NA	NA	24.5	5.03	09/15/2025	09:19
14	Q3096-01	1	Water	NA	NA	20.6	5.98	09/15/2025	09:30
15	Q3096-01DUP	1	Water	NA	NA	20.9	5.99	09/15/2025	09:33
16	CCV2	1	Water	NA	NA	20.2	12.02	09/15/2025	09:35

WORKLIST(Hardcopy Internal Chain)

UB 138185

WorkList Name : ph w q3096 WorkList ID : 191854 Department : Wet-Chemistry Date : 09-15-2025 07:52:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3081-01	DRAIN-WATER-TANK-1	Water	pH	Cool 4 deg C	MAJO01	D11	09/11/2025	9040C
Q3088-02	90525-LQ	Water	pH	Cool 4 deg C	PSEG03	D31	09/12/2025	9040C
Q3090-01	90825-B	Water	pH	Cool 4 deg C	PSEG03	D41	09/12/2025	9040C
Q3093-01	MH-9-12-25	Water	pH	Cool 4 deg C	EURO03	D31	09/12/2025	9040C
Q3096-01	DRAIN-WATER-TANK-1	Water	pH	Cool 4 deg C	MAJO01	D31	09/12/2025	9040C

Date/Time 09-15-25 08:10

Raw Sample Received by: JB (e2c)

Raw Sample Relinquished by: JWS

Date/Time 09-15-25

Raw Sample Received by: JWS

Raw Sample Relinquished by: JB