

Initial and Continuing Calibration Verification

Client: Major Products

SDG No.: Q3081

Project: Floor Drain Water 2025

RunNo.: LB137185

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.00	7	100	90-110	09/15/2025
Sample ID:	CCV1						
pH		pH	2.02	2.00	101	90-110	09/15/2025
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	09/15/2025