

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3083

Client ID : SSTDCCC0.4

Date Analyzed: 09/19/2025

Lab File ID: BN037813.D

Time Analyzed: 14:19

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	4812	7.833	12410	10.64	5585	14.48
UPPER LIMIT	9624	8.333	24820	11.137	11170	14.984
LOWER LIMIT	2406	7.333	6205	10.137	2792.5	13.984
EPA SAMPLE NO.						
01 MW3SDL	5937	7.83	13519	10.64	6156	14.48

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.