

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3083

Client ID: SSTDCCC0.4

Date Analyzed: 09/18/2025

Lab File ID: BN037776.D

Time Analyzed: 14:32

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	14526	17.235	10438	21.411	10600	23.744
UPPER LIMIT	29052	17.735	20876	21.911	21200	24.244
LOWER LIMIT	7263	16.735	5219	20.911	5300	23.244
EPA SAMPLE NO.						
01 PB169693BL	11652	17.24	9221	21.41	9002	23.75
02 PB169693BS	12386	17.24	10095	21.41	10200	23.74
03 MW3S	17415	17.24	17835	21.41	19008	23.74

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.