

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3083

Client ID: SSTDCCC0.4

Date Analyzed: 09/19/2025

Lab File ID: BN037813.D

Time Analyzed: 14:19

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

01

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	9638	17.235	8466	21.411	8084	23.739
UPPER LIMIT	19276	17.735	16932	21.911	16168	24.239
LOWER LIMIT	4819	16.735	4233	20.911	4042	23.239
EPA SAMPLE NO.						
MW3SDL	11729	17.24	10935	21.41	10906	23.74

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.