

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3083

Client ID : SSTDCCC040

Date Analyzed: 09/17/2025

Lab File ID: BP025769.D

Time Analyzed: 12:03

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	173185	7.755	666754	10.53	429441	14.37
UPPER LIMIT	346370	8.255	1333510	11.025	858882	14.872
LOWER LIMIT	86592.5	7.255	333377	10.025	214721	13.872
EPA SAMPLE NO.						
01 PB169719BL	179386	7.75	669415	10.53	415180	14.37
02 PB169719BS	164300	7.76	643426	10.52	416776	14.37
03 PB169719BSD	164703	7.76	637233	10.53	413759	14.37

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.