

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN082125

Review By	John Carlone	Review On	8/22/2025 9:17:52 AM
Supervise By	Maresh Dadoda	Supervise On	8/25/2025 8:11:38 AM
SubDirectory	VN082125	HP Acquire Method	HP Processing Method 82N082125W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135214 VP135215,VP135216,VP135217,VP135218,VP135219,VP135220 VP135221

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087614.D	21 Aug 2025 09:30		JC\MD	Ok
2	VSTDCCC020	VSTDCCC020	VN087615.D	21 Aug 2025 10:04	Need ICAL	JC\MD	Not Ok
3	VN0821WBS01	VN0821WBS01	VN087616.D	21 Aug 2025 10:37		JC\MD	Not Ok
4	VN0821WBSD01	VN0821WBSD01	VN087617.D	21 Aug 2025 11:11		JC\MD	Not Ok
5	VN0821WBL01	VN0821WBL01	VN087618.D	21 Aug 2025 11:36		JC\MD	Not Ok
6	VSTDICC001	VSTDICC001	VN087619.D	21 Aug 2025 12:09	LR-10,20	JC\MD	Ok,M
7	VSTDICC005	VSTDICC005	VN087620.D	21 Aug 2025 13:08		JC\MD	Ok,M
8	VSTDICC020	VSTDICC020	VN087621.D	21 Aug 2025 13:41	method not used for #N-amyle Acetate	JC\MD	Ok,M
9	VSTDICCC050	VSTDICCC050	VN087622.D	21 Aug 2025 14:02		JC\MD	Ok,M
10	VSTDICC100	VSTDICC100	VN087623.D	21 Aug 2025 14:23		JC\MD	Ok,M
11	VSTDICC150	VSTDICC150	VN087624.D	21 Aug 2025 14:44		JC\MD	Ok,M
12	IBLK	IBLK	VN087625.D	21 Aug 2025 15:05		JC\MD	Ok
13	VSTDICV050	ICVVN082125	VN087626.D	21 Aug 2025 15:47	Icv Failed for Com.# 64 for DOD	JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN091125

Review By	Mahesh Dadoda	Review On	9/12/2025 1:48:33 PM				
Supervise By	Semsettin Yesilyurt	Supervise On	9/12/2025 1:52:34 PM				
SubDirectory	VN091125	HP Acquire Method	MSVOA_N	HP Processing Method	VN082125W.M		
STD. NAME		STD REF.#					
Tune/Reschk Initial Calibration Stds		VP135417					
CCC		VP135415,VP135416					
Internal Standard/PEM		VP134956					
ICV/I.BLK							
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087791.D	11 Sep 2025 09:53		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN087792.D	11 Sep 2025 12:49		JC\MD	Ok,M
3	VN0911MBL01	VN0911MBL01	VN087793.D	11 Sep 2025 13:25		JC\MD	Ok
4	VN0911WBL01	VN0911WBL01	VN087794.D	11 Sep 2025 13:45		JC\MD	Ok
5	VN0911WBS01	VN0911WBS01	VN087795.D	11 Sep 2025 14:21	BS Failed High For Com.#40,65,68	JC\MD	Ok,M
6	VN0911WBSD01	VN0911WBSD01	VN087796.D	11 Sep 2025 14:42	BSD Failed High For Com.#39,40,65,67,68,69	JC\MD	Ok,M
7	Q3078-01	250910074-01-VOA	VN087797.D	11 Sep 2025 15:03	Need 10X;BS,BSD Failed High,vial A pH#8	JC\MD	Dilution
8	VIBLK	VIBLK	VN087798.D	11 Sep 2025 15:25		JC\MD	Ok
9	Q3078-02	250910064-01-TRIP-BL	VN087799.D	11 Sep 2025 15:45	TB,vial A pH#7	JC\MD	Ok
10	PB169644TB	PB169644TB	VN087800.D	11 Sep 2025 16:06	vial A pH# 5	JC\MD	Ok,M
11	Q3054-04	MH-385	VN087801.D	11 Sep 2025 16:27	vial A pH# 5	JC\MD	Ok,M
12	Q3054-08	MH-384	VN087802.D	11 Sep 2025 16:48	vial B pH# 5	JC\MD	Ok
13	Q3072-02	30-GARFIELD-PL-COM	VN087803.D	11 Sep 2025 17:09	vial A pH# 5	JC\MD	Ok,M
14	Q3083-01	MW4S	VN087804.D	11 Sep 2025 17:30	BS failed High;BSD failed High,vial A pH#<2	JC\MD	ReRun
15	Q3083-02	MW4I	VN087805.D	11 Sep 2025 17:51	vial A pH#<3	JC\MD	Ok
16	Q3083-03	MW3S	VN087806.D	11 Sep 2025 18:11	vial A pH# <2	JC\MD	Ok
17	Q3078-01DL	250910074-01-VOADL	VN087807.D	11 Sep 2025 18:32	BS,BSD Failed High,vial C pH# 8,NO SAMPLE TO RERUN	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN091125

Review By	Mahesh Dadoda	Review On	9/12/2025 1:48:33 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	9/12/2025 1:52:34 PM		
SubDirectory	VN091125	HP Acquire Method	MSVOA_N	HP Processing Method	VN082125W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135417
CCC	VP135415,VP135416
Internal Standard/PEM	VP134956
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	VN0911WBS02	VN0911WBS02	VN087808.D	11 Sep 2025 18:53	Surrogate Fail	JC\MD	Not Ok
19	VSTDCCC050	VSTDCCC050EC	VN087809.D	11 Sep 2025 19:14		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN091525

Review By	Mahesh Dadoda	Review On	9/16/2025 7:59:53 AM
Supervise By	Semsettin Yesilyurt	Supervise On	9/16/2025 8:03:28 AM
SubDirectory	VN091525	HP Acquire Method	HP Processing Method 82N082125W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135447
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135448,VP135449

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087819.D	15 Sep 2025 08:18		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN087820.D	15 Sep 2025 08:39	pH#Lot#V12668	JC\MD	Ok,M
3	VN0915MBL01	VN0915MBL01	VN087821.D	15 Sep 2025 09:14		JC\MD	Ok
4	VN0915WBL01	VN0915WBL01	VN087822.D	15 Sep 2025 09:35		JC\MD	Ok
5	VN0915WBS01	VN0915WBS01	VN087823.D	15 Sep 2025 09:56		JC\MD	Ok,M
6	VN0915WBSD01	VN0915WBSD01	VN087824.D	15 Sep 2025 10:29		JC\MD	Ok,M
7	Q3083-01	MW4S	VN087825.D	15 Sep 2025 10:50	vial B pH<2	JC\MD	Ok
8	Q3088-02	90525-LQ	VN087826.D	15 Sep 2025 11:11	vial A pH<2 brown/ turbid sample	JC\MD	Ok
9	PB169678TB	PB169678TB	VN087827.D	15 Sep 2025 11:32		JC\MD	Ok
10	Q3082-02	WC-29	VN087828.D	15 Sep 2025 11:53	vial A pH#5.0	JC\MD	Ok
11	Q3082-04	WC-30	VN087829.D	15 Sep 2025 12:14	vial A pH#5.0	JC\MD	Ok,M
12	Q3082-06	WC-31	VN087830.D	15 Sep 2025 12:35	vial A pH#5.0	JC\MD	Ok
13	Q3082-08	WC-32	VN087831.D	15 Sep 2025 12:56	vial A pH#5.0	JC\MD	Ok
14	Q3082-10	WC-33	VN087832.D	15 Sep 2025 13:17	vial A pH#5.0	JC\MD	Ok,M
15	Q3092-01	291376	VN087833.D	15 Sep 2025 13:38	vial A pH#5.0	JC\MD	Ok
16	Q3092-02	291376	VN087834.D	15 Sep 2025 13:59	vial A pH#5.0	JC\MD	Ok
17	Q3092-03	279182	VN087835.D	15 Sep 2025 14:20	vial A pH#5.0	JC\MD	Ok,M

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN091525

Review By	Maresh Dadoda	Review On	9/16/2025 7:59:53 AM
Supervise By	Semsettin Yesilyurt	Supervise On	9/16/2025 8:03:28 AM
SubDirectory	VN091525	HP Acquire Method	HP Processing Method 82N082125W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135447		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135448,VP135449		

18	Q3092-04	279182	VN087836.D	15 Sep 2025 14:41	vial A pH#5.0	JC\MD	Ok
19	VN0915MBS01	VN0915MBS01	VN087837.D	15 Sep 2025 15:02		JC\MD	Ok,M
20	VN0915MBSD01	VN0915MBSD01	VN087838.D	15 Sep 2025 15:22		JC\MD	Ok,M
21	Q3058-01ME	RPXY5ME	VN087839.D	15 Sep 2025 15:43		JC\MD	Ok
22	Q3075-04	MOO-25-0258	VN087840.D	15 Sep 2025 16:04	cloth material	JC\MD	Ok
23	Q3074-01	MOO-25-0141	VN087841.D	15 Sep 2025 16:25	cloth material	JC\MD	Ok
24	Q3073-02	BEL-25-0023	VN087842.D	15 Sep 2025 16:46	cloth material	JC\MD	Ok
25	Q3075-06	MOO-25-0259	VN087843.D	15 Sep 2025 17:06	cloth material	JC\MD	Ok
26	Q3077-01	LAW-25-0137	VN087844.D	15 Sep 2025 17:27	cloth material run straight MEOH	JC\MD	Not Ok
27	VSTDCCC050	VSTDCCC050EC	VN087845.D	15 Sep 2025 18:09		JC\MD	Ok,M

M : Manual Integration