

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	Q3083	Analytical Method:	SW8260-Low
Client:	G Environmental	Datafile :	VN087796.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0911WBSD01	Dichlorodifluoromethane	20	22.3	ug/L	112	2		40 (69)	160 (116)	20 (19)
	Chloromethane	20	22.2	ug/L	111	10		40 (65)	160 (116)	20 (21)
	Vinyl chloride	20	21.8	ug/L	109	2		70 (65)	130 (117)	20 (19)
	Bromomethane	20	24.0	ug/L	120	8		40 (58)	160 (125)	20 (20)
	Chloroethane	20	20.2	ug/L	101	5		40 (56)	160 (128)	20 (20)
	Trichlorofluoromethane	20	22.8	ug/L	114	2		40 (73)	160 (115)	20 (16)
	1,1,2-Trichlorotrifluoroethane	20	22.0	ug/L	110	2		70 (80)	130 (112)	20 (15)
	1,1-Dichloroethene	20	21.3	ug/L	106	1		70 (74)	130 (110)	20 (20)
	Acetone	100	88.8	ug/L	89	8		40 (60)	160 (125)	20 (20)
	Carbon disulfide	20	21.7	ug/L	109	3		40 (64)	160 (112)	20 (20)
	Methyl tert-butyl Ether	20	19.6	ug/L	98	5		70 (78)	130 (114)	20 (20)
	Methyl Acetate	20	19.0	ug/L	95	7		70 (67)	130 (125)	20 (20)
	Methylene Chloride	20	20.5	ug/L	103	9		70 (72)	130 (114)	20 (20)
	trans-1,2-Dichloroethene	20	20.0	ug/L	100	8		70 (75)	130 (108)	20 (16)
	1,1-Dichloroethane	20	19.5	ug/L	98	6		70 (78)	130 (112)	20 (20)
	Cyclohexane	20	21.0	ug/L	105	1		70 (75)	130 (110)	20 (20)
	2-Butanone	100	97.3	ug/L	97	3		40 (65)	160 (122)	20 (26)
	Carbon Tetrachloride	20	23.7	ug/L	119	5		70 (77)	130 (113)	20 (15)
	cis-1,2-Dichloroethene	20	19.4	ug/L	97	8		70 (77)	130 (110)	20 (20)
	Bromochloromethane	20	19.4	ug/L	97	6		70 (70)	130 (124)	20 (20)
	Chloroform	20	20.4	ug/L	102	6		70 (79)	130 (113)	20 (20)
	1,1,1-Trichloroethane	20	20.6	ug/L	103	3		70 (80)	130 (108)	20 (20)
	Methylcyclohexane	20	23.4	ug/L	117	10		70 (72)	130 (115)	20 (20)
	Benzene	20	22.0	ug/L	110	1		70 (82)	130 (109)	20 (15)
	1,2-Dichloroethane	20	21.3	ug/L	106	2		70 (80)	130 (115)	20 (20)
	Trichloroethene	20	22.4	ug/L	112	2		70 (77)	130 (113)	20 (15)
	1,2-Dichloropropane	20	20.7	ug/L	104	3		70 (83)	130 (111)	20 (16)
	Bromodichloromethane	20	22.2	ug/L	111	1		70 (83)	130 (110)	20 (16)
	4-Methyl-2-Pentanone	100	110	ug/L	110	0		40 (74)	160 (118)	20 (25)
	Toluene	20	22.0	ug/L	110	3		70 (82)	130 (110)	20 (16)
	t-1,3-Dichloropropene	20	21.4	ug/L	107	4		70 (79)	130 (110)	20 (20)
	cis-1,3-Dichloropropene	20	21.4	ug/L	107	5		70 (82)	130 (110)	20 (16)
	1,1,2-Trichloroethane	20	22.4	ug/L	112	1		70 (83)	130 (112)	20 (20)
	2-Hexanone	100	110	ug/L	110	0		40 (73)	160 (117)	20 (25)
	Dibromochloromethane	20	22.1	ug/L	111	0		70 (82)	130 (110)	20 (20)
	1,2-Dibromoethane	20	22.1	ug/L	111	1		70 (81)	130 (110)	20 (20)
	Tetrachloroethene	20	22.9	ug/L	115	4		70 (67)	130 (123)	20 (15)
	Chlorobenzene	20	21.9	ug/L	110	3		70 (82)	130 (109)	20 (15)
	Ethyl Benzene	20	22.3	ug/L	112	4		70 (83)	130 (109)	20 (16)
	m/p-Xylenes	40	46.3	ug/L	116	3		70 (82)	130 (110)	20 (15)
	o-Xylene	20	22.1	ug/L	111	2		70 (83)	130 (109)	20 (20)
	Styrene	20	22.6	ug/L	113	1		70 (80)	130 (111)	20 (17)
	Bromoform	20	24.1	ug/L	121	0		70 (79)	130 (109)	20 (20)
	Isopropylbenzene	20	20.8	ug/L	104	1		70 (83)	130 (112)	20 (29)
	1,1,2,2-Tetrachloroethane	20	21.4	ug/L	107	4		70 (76)	130 (118)	20 (20)
	1,3-Dichlorobenzene	20	21.4	ug/L	107	3		70 (82)	130 (108)	20 (20)
	1,4-Dichlorobenzene	20	20.7	ug/L	104	2		70 (82)	130 (107)	20 (15)
	1,2-Dichlorobenzene	20	21.9	ug/L	110	0		70 (82)	130 (109)	20 (20)

() = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3083</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>G Environmental</u>	Datafile :	<u>VN087796.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0911WBSD01	1,2-Dibromo-3-Chloropropane	20	19.8	ug/L	99	1		40 (68)	160 (112)	20 (20)
	1,2,4-Trichlorobenzene	20	20.7	ug/L	104	4		70 (75)	130 (113)	20 (29)
	1,2,3-Trichlorobenzene	20	21.6	ug/L	108	0		70 (76)	130 (114)	20 (29)