

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q3083
 Lab File ID: VN087792.D Date Analyzed: 09/11/2025
 Instrument ID: MSVOA_N Time Analyzed: 12:49
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	284150	8.21	535670	9.08	504865	11.85
UPPER LIMIT	568300	8.706	1071340	9.582	1009730	12.347
LOWER LIMIT	142075	7.706	267835	8.582	252433	11.347
EPA SAMPLE NO.						
MW4I	220322	8.21	503440	9.08	493455	11.84
MW3S	165866	8.21	379915	9.08	368148	11.84
VN0911WBL01	217125	8.21	505649	9.08	478608	11.84
VN0911WBS01	264571	8.21	504460	9.08	466377	11.84
VN0911WBSD01	284367	8.21	516573	9.08	469236	11.84

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.