

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>GENV01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3083</u>
Lab File ID:	<u>VN087792.D</u>	Date Analyzed:	<u>09/11/2025</u>
Instrument ID:	<u>MSVOA_N</u>	Time Analyzed:	<u>12:49</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	262340	13.77				
UPPER LIMIT	524680	14.27				
LOWER LIMIT	131170	13.27				
EPA SAMPLE NO.						
MW4I	228190	13.77				
MW3S	175104	13.77				
VN0911WBL01	224924	13.77				
VN0911WBS01	243902	13.77				
VN0911WBSD01	252435	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.