

## FORM 16-IN

## INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: ARC Sayreville MA No:  SDG No.: Q3084  
Instrument ID: CV1 Start Date : 09/24/2025  
Analytical Method: CVAA Run Batch : LB137291

| Analyte | Corr. Coeff. | Slope    | Intercept | Calib. Type | Weighting |
|---------|--------------|----------|-----------|-------------|-----------|
| Mercury | 0.998000     | 0.000365 | -0.022200 | Lin. Reg    | NONE      |