

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: ARC Sayrevill MA No. : S DG No.: Q3084Initial Calibration Verification Source : MP87328Continuing Calibration Verification Source : MP87330Run Batch: LB137291 Analytical Method: CVAAConcentration Units: $\mu\text{g/L}$

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV44				ID: CCV33				ID: CCV34		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury	4.0	3.8	96		5.0	5.1	102		5.1	102	

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: ARC Sayrevill MA No. : S DG No.: Q3084
Initial Calibration Verification Source : MP87328
Continuing Calibration Verification Source : MP87330
Run Batch: LB137291 Analytical Method: CVAA
Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV35				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury					5.0	5.1	102				