

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: ARC Sayrevill MA No. : SDG No.: Q3084
 Initial Calibration Verification Source : MP87085
 Continuing Calibration Verification Source : MP87086
 Run Batch: LB137284 Analytical Method: ICP-MS
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV001				ID: CCV001				ID: CCV002		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum	1600	1600	98	2	50000	51000	101	1	52000	103	1
Antimony	400	420	105	2	500	500	101	6	520	105	1
Arsenic	400	410	101	2	500	490	98	1	490	98	0
Beryllium	40.0	43	107	4	500	500	100	6	510	102	3
Cadmium	200	210	104	3	500	490	98	6	500	100	2
Chromium	160	160	100	1	500	500	100	1	500	100	1
Iron	800	830	104	2	125000	130000	105	0	130000	105	2
Lead	400	390	97	2	2500	2500	100	5	2500	100	1
Manganese	400	360	91	2	5000	5000	101	1	5000	101	2
Nickel	400	410	102	2	500	490	98	1	490	97	1
Thallium	400	380	96	2	500	500	100	5	510	103	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: ARC Sayrevill MA No. : S DG No.: Q3084Initial Calibration Verification Source : MP87085Continuing Calibration Verification Source : MP87086Run Batch: LB137284 Analytical Method: ICP-MSConcentration Units: $\mu\text{g/L}$

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV003				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					50000	51000	102	2			
Antimony					500	520	103	1			
Arsenic					500	480	95	1			
Beryllium					500	520	104	1			
Cadmium					500	490	99	1			
Chromium					500	500	99	1			
Iron					125000	130000	106	2			
Lead					2500	2500	102	2			
Manganese					5000	5000	101	2			
Nickel					500	480	97	2			
Thallium					500	520	104	1			