

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091625\
 Data File : BN037732.D
 Acq On : 16 Sep 2025 14:48
 Operator : RC/JU
 Sample : PB169690BS
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS690

Quant Time: Sep 16 16:02:49 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN091225.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 12 15:43:05 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.840	152	171200	20.000	ng/u1	0.00
20) Naphthalene-d8	10.646	136	626755	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.492	164	353199	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.239	188	636770	20.000	ng/u1	0.00
79) Chrysene-d12	21.492	240	659268	20.000	ng/u1	0.00
88) Perylene-d12	24.574	264	693425	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.275	96	33145	8.201	ng/uL	0.01
4) Pyridine-d5	3.681	84	473146	42.098	ng/u1	0.00
7) Phenol-d5	6.987	99	502746	39.360	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.163	67	315058	40.004	ng/u1	0.00
11) 2-Chlorophenol-d4	7.369	132	441317	41.867	ng/u1	0.00
15) 4-Methylphenol-d8	8.540	113	399552	39.738	ng/u1	0.00
21) Nitrobenzene-d5	8.993	128	204239	43.210	ng/u1	0.00
24) 2-Nitrophenol-d4	9.722	143	173341	44.824	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.257	165	387017	43.827	ng/u1	0.00
31) 4-Chloroaniline-d4	10.769	131	519555	38.787	ng/u1	0.00
46) Dimethylphthalate-d6	13.904	166	1015414	43.057	ng/u1	0.00
49) Acenaphthylene-d8	14.187	160	1283757	43.563	ng/u1	0.00
54) 4-Nitrophenol-d4	14.675	143	191516	43.333	ng/u1	0.00
60) Fluorene-d10	15.486	176	914645	43.363	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.592	200	106216	36.998	ng/u1	0.00
73) Anthracene-d10	17.339	188	1370075	43.275	ng/u1	0.00
81) Pyrene-d10	19.627	212	1468228	42.604	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.368	264	1627953	46.499	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.311	88	62353	14.491	ng/uL	97
5) Pyridine	3.705	79	426468	36.722	ng/u1	94
6) Benzaldehyde	6.969	77	49845	8.330	ng/u1	94
8) Phenol	7.016	94	488806	36.772	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.258	93	397957	37.548	ng/u1	98
12) 2-Chlorophenol	7.399	128	423076	38.378	ng/u1	98
13) 2-Methylphenol	8.275	108	357365	36.463	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.363	45	518502	35.516	ng/u1	99
16) Acetophenone	8.657	105	589427	36.801	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.646	70	270671	37.081	ng/u1	96
18) 4-Methylphenol	8.604	108	386931	36.595	ng/u1	99
19) Hexachloroethane	8.928	117	181115	38.586	ng/u1	98
22) Nitrobenzene	9.034	77	421810	39.000	ng/u1	96
23) Isophorone	9.569	82	715229	38.112	ng/u1	99
25) 2-Nitrophenol	9.751	139	189233	41.485	ng/u1	98
26) 2,4-Dimethylphenol	9.810	107	407181	39.123	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.051	93	485040	38.778	ng/u1	99
29) 2,4-Dichlorophenol	10.287	162	353245	39.200	ng/u1	97
30) Naphthalene	10.693	128	1314476	39.266	ng/u1	99
32) 4-Chloroaniline	10.793	127	280951	20.975	ng/u1	97
33) Hexachlorobutadiene	10.993	225	252931	40.253	ng/u1	97
34) Caprolactam	11.551	113	94030	35.222	ng/u1	94
35) 4-Chloro-3-methylphenol	11.922	107	336672	39.672	ng/u1	99
36) 2-Methylnaphthalene	12.310	142	876096	39.798	ng/u1	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091625\
 Data File : BN037732.D
 Acq On : 16 Sep 2025 14:48
 Operator : RC/JU
 Sample : PB169690BS
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS690

Quant Time: Sep 16 16:02:49 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN091225.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 12 15:43:05 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.534	142	851011	39.092	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.681	216	425768	38.646	ng/ul	98
40) Hexachlorocyclopentadiene	12.669	237	248447	35.177	ng/ul	99
41) 2,4,6-Trichlorophenol	12.916	196	245153	41.133	ng/ul	97
42) 2,4,5-Trichlorophenol	12.987	196	281447	40.699	ng/ul	99
43) 1,1'-Biphenyl	13.328	154	1059483	38.951	ng/ul	99
44) 2-Chloronaphthalene	13.363	162	835788	39.031	ng/ul	98
45) 2-Nitroaniline	13.557	65	194563	39.896	ng/ul	97
47) Dimethylphthalate	13.951	163	942164	39.953	ng/ul	100
48) 2,6-Dinitrotoluene	14.057	165	173279	41.098	ng/ul	98
50) Acenaphthylene	14.216	152	1360428	39.781	ng/ul	99
51) 3-Nitroaniline	14.387	138	189997	42.046	ng/ul	95
52) Acenaphthene	14.557	153	881281	39.602	ng/ul	97
53) 2,4-Dinitrophenol	14.587	184	58506	32.832	ng/ul	98
55) 4-Nitrophenol	14.687	109	135118	40.145	ng/ul	97
56) Dibenzofuran	14.892	168	1191439	39.242	ng/ul	99
57) 2,4-Dinitrotoluene	14.845	165	262783	43.530	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.116	232	213828	44.211	ng/ul	99
59) Diethylphthalate	15.322	149	906406	40.453	ng/ul	98
61) Fluorene	15.545	166	977499	39.725	ng/ul	95
62) 4-Chlorophenyl-phenyle...	15.539	204	462848	39.094	ng/ul	96
63) 4-Nitroaniline	15.545	138	220438	51.851	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.610	198	112042	34.938	ng/ul	98
67) N-Nitrosodiphenylamine	15.751	169	812094	39.762	ng/ul	99
68) 4-Bromophenyl-phenylether	16.433	248	290112	40.511	ng/ul	94
69) Hexachlorobenzene	16.551	284	346296	39.509	ng/ul	100
70) Atrazine	16.704	200	255750	37.265	ng/ul	99
71) Pentachlorophenol	16.886	266	184227	38.289	ng/ul	95
72) Phenanthrene	17.281	178	1463612	39.666	ng/ul	98
74) Anthracene	17.375	178	1492446	39.890	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.292	216	414426	36.581	ng/ul	96
76) Pentachlorobenzene	14.816	250	407113	37.515	ng/ul	98
77) Carbazole	17.639	167	1332321	40.197	ng/ul	99
78) Di-n-butylphthalate	18.222	149	1458400	43.362	ng/ul	100
80) Fluoranthene	19.298	202	1605399	39.920	ng/ul	99
82) Pyrene	19.657	202	1697432	39.686	ng/ul	99
83) Butylbenzylphthalate	20.563	149	539050	44.640	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.392	252	479575	37.664	ng/ul	95
85) Benzo(a)anthracene	21.469	228	1656640	39.181	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.416	149	833440	45.785	ng/ul	99
87) Chrysene	21.533	228	1583444	39.166	ng/ul	98
89) Di-n-octyl phthalate	22.586	149	1239740	37.959	ng/ul	100
90) Benzo(b)fluoranthene	23.598	252	1695413	41.416	ng/ul	98
91) Benzo(k)fluoranthene	23.668	252	1692754	40.980	ng/ul	97
93) Benzo(a)pyrene	24.433	252	1639290	41.778	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	28.068	276	2259547	44.009	ng/ul	99
95) Dibenzo(a,h)anthracene	28.139	278	1874585	44.173	ng/ul	98
96) Benzo(g,h,i)perylene	29.150	276	1838111	43.691	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091625\
Data File : BN037732.D
Acq On : 16 Sep 2025 14:48
Operator : RC/JU
Sample : PB169690BS
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SLCS690

Quant Time: Sep 16 16:02:49 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN091225.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Sep 12 15:43:05 2025
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 14.491 ng/uL
 RT: 3.311 min Scan# 31
 Delta R.T. 0.006 min
 Lab File: BN037732.D
 Acq: 16 Sep 2025 14:48




Instrument :
 BNA_N
 ClientSampleId :
 SLCS690


Tgt Ion: 88 Resp: 62353
 Ion Ratio Lower Upper
 88 100
 43 27.7 23.0 34.6
 58 72.7 60.3 90.5



#5
 Pyridine
 Concen: 36.722 ng/uL
 RT: 3.705 min Scan# 105
 Delta R.T. 0.012 min
 Lab File: BN037732.D
 Acq: 16 Sep 2025 14:48

Tgt Ion: 79 Resp: 426468
 Ion Ratio Lower Upper
 79 100
 52 68.9 59.8 89.8
 51 34.4 29.2 43.8









#6
Benzaldehyde
Concen: 8.330 ng/ul
RT: 6.969 min Scan# 660
Delta R.T. 0.006 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SLCS690


Tgt Ion: 77 Resp: 49845
Ion Ratio Lower Upper
77 100
105 105.3 77.5 116.3
106 96.4 75.0 112.4



#8
Phenol
Concen: 36.772 ng/ul
RT: 7.016 min Scan# 668
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48

Tgt Ion: 94 Resp: 488806
Ion Ratio Lower Upper
94 100
65 27.4 22.1 33.1
66 35.8 28.8 43.2









#10
 Bis(2-Chloroethyl)ether
 Concen: 37.548 ng/ul
 RT: 7.258 min Scan# 709
 Delta R.T. 0.006 min
 Lab File: BN037732.D
 Acq: 16 Sep 2025 14:48




Instrument :
 BNA_N
 ClientSampleId :
 SLCS690


Tgt Ion: 93 Resp: 397957
 Ion Ratio Lower Upper
 93 100
 63 73.2 59.7 89.5
 95 33.3 25.1 37.7



#12
 2-Chlorophenol
 Concen: 38.378 ng/ul
 RT: 7.399 min Scan# 733
 Delta R.T. -0.000 min
 Lab File: BN037732.D
 Acq: 16 Sep 2025 14:48

Tgt Ion:128 Resp: 423076
 Ion Ratio Lower Upper
 128 100
 64 42.6 33.6 50.4
 130 32.7 24.7 37.1









#13
2-Methylphenol
Concen: 36.463 ng/ul
RT: 8.275 min Scan# 882
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SLCS690


Tgt Ion:108 Resp: 357365
Ion Ratio Lower Upper
108 100
107 87.5 69.4 104.2



#14
2,2'-oxybis(1-Chloropropane)
Concen: 35.516 ng/ul
RT: 8.363 min Scan# 897
Delta R.T. -0.006 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48

Tgt Ion: 45 Resp: 518502
Ion Ratio Lower Upper
45 100
77 15.1 11.6 17.4
79 10.9 8.3 12.5









#16
Acetophenone
Concen: 36.801 ng/ul
RT: 8.657 min Scan# 947
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SLCS690


Tgt Ion:105 Resp: 589427
Ion Ratio Lower Upper
105 100
77 72.8 58.9 88.3
51 24.4 21.1 31.7



#17
N-Nitroso-di-n-propylamine
Concen: 37.081 ng/ul
RT: 8.646 min Scan# 945
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48

Tgt Ion: 70 Resp: 270671
Ion Ratio Lower Upper
70 100
42 56.2 47.8 71.8
101 10.1 7.9 11.9
130 23.2 19.4 29.0









#18
4-Methylphenol
Concen: 36.595 ng/ul
RT: 8.604 min Scan# 911
Delta R.T. 0.006 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SLCS690


Tgt Ion:108 Resp: 386931
Ion Ratio Lower Upper
108 100
107 112.0 88.4 132.6



#19
Hexachloroethane
Concen: 38.586 ng/ul
RT: 8.928 min Scan# 993
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48

Tgt Ion:117 Resp: 181115
Ion Ratio Lower Upper
117 100
201 98.1 78.1 117.1
199 62.8 48.1 72.1





#22

Nitrobenzene

Concen: 39.000 ng/ul

RT: 9.034 min Scan# 1101

Delta R.T. -0.000 min

Lab File: BN037732.D

Acq: 16 Sep 2025 14:48

Instrument :

BNA_N

ClientSampleId :

SLCS690





Tgt Ion: 77 Resp: 421810

Ion Ratio Lower Upper

77 100

123 50.0 37.5 56.3

65 13.9 10.5 15.7



#23

Isophorone

Concen: 38.112 ng/ul

RT: 9.569 min Scan# 1102

Delta R.T. -0.000 min

Lab File: BN037732.D

Acq: 16 Sep 2025 14:48




Tgt Ion: 82 Resp: 715229


Ion Ratio Lower Upper

82 100

95 7.4 5.5 8.3

138 19.4 15.0 22.6









#25
2-Nitrophenol
Concen: 41.485 ng/ul
RT: 9.751 min Scan# 1133
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SLCS690


Tgt Ion:139 Resp: 189233
Ion Ratio Lower Upper
139 100
65 48.1 37.9 56.9
109 30.0 23.0 34.4



#26
2,4-Dimethylphenol
Concen: 39.123 ng/ul
RT: 9.810 min Scan# 1143
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48

Tgt Ion:107 Resp: 407181
Ion Ratio Lower Upper
107 100
121 51.2 41.4 62.2
122 89.0 69.9 104.9









#27
 Bis(2-Chloroethoxy)methane
 Concen: 38.778 ng/ul
 RT: 10.051 min Scan# 1184
 Delta R.T. -0.000 min
 Lab File: BN037732.D
 Acq: 16 Sep 2025 14:48




Instrument :
 BNA_N
 ClientSampleId :
 SLCS690


Tgt Ion: 93 Resp: 485040
 Ion Ratio Lower Upper
 93 100
 95 33.3 26.0 39.0
 123 11.6 9.5 14.3



#29
 2,4-Dichlorophenol
 Concen: 39.200 ng/ul
 RT: 10.287 min Scan# 1224
 Delta R.T. -0.000 min
 Lab File: BN037732.D
 Acq: 16 Sep 2025 14:48

Tgt Ion:162 Resp: 353245
 Ion Ratio Lower Upper
 162 100
 164 67.0 52.7 79.1
 98 35.8 26.3 39.5









#30
 Naphthalene
 Concen: 39.266 ng/ul
 RT: 10.693 min Scan# 1293
 Delta R.T. -0.000 min
 Lab File: BN037732.D
 Acq: 16 Sep 2025 14:48

Instrument :
 BNA_N
 ClientSampleId :
 SLCS690

Tgt Ion:128 Resp: 1314476




Ion	Ratio	Lower	Upper
128	100		
129	11.0	9.1	13.7
127	13.1	10.9	16.3




#32
 4-Chloroaniline
 Concen: 20.975 ng/ul
 RT: 10.793 min Scan# 1310
 Delta R.T. -0.006 min
 Lab File: BN037732.D
 Acq: 16 Sep 2025 14:48

Tgt Ion:127 Resp: 280951

Ion	Ratio	Lower	Upper
127	100		
129	33.3	25.4	38.0





#33

Hexachlorobutadiene

Concen: 40.253 ng/ul

RT: 10.993 min Scan# 11

Delta R.T. -0.000 min

Lab File: BN037732.D

Acq: 16 Sep 2025 14:48

Instrument :

BNA_N

ClientSampleId :

SLCS690





Tgt Ion:225 Resp: 252931

Ion Ratio Lower Upper

225 100

223 60.3 50.3 75.5

227 64.3 49.4 74.0



#34

Caprolactam

Concen: 35.222 ng/ul

RT: 11.551 min Scan# 1439

Delta R.T. -0.000 min

Lab File: BN037732.D

Acq: 16 Sep 2025 14:48




Tgt Ion:113 Resp: 94030


Ion Ratio Lower Upper

113 100

55 151.1 126.6 190.0

56 112.0 96.4 144.6









#35
4-Chloro-3-methylphenol
Concen: 39.672 ng/ul
RT: 11.922 min Scan# 1502
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SLCS690


Tgt Ion:107 Resp: 336672
Ion Ratio Lower Upper
107 100
144 28.1 21.8 32.8
142 83.0 66.6 100.0



#36
2-Methylnaphthalene
Concen: 39.798 ng/ul
RT: 12.310 min Scan# 1568
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48

Tgt Ion:142 Resp: 876096
Ion Ratio Lower Upper
142 100
141 90.9 69.8 104.8









#37
1-Methylnaphthalene
Concen: 39.092 ng/ul
RT: 12.534 min Scan# 1606
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SLCS690


Tgt Ion:142 Resp: 851011
Ion Ratio Lower Upper
142 100
141 91.8 74.6 111.8
116 3.4 2.9 4.3



#39
1,2,4,5-Tetrachlorobenzene
Concen: 38.646 ng/ul
RT: 12.681 min Scan# 1631
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48

Tgt Ion:216 Resp: 425768
Ion Ratio Lower Upper
216 100
214 80.7 65.5 98.3
179 21.3 17.4 26.2
108 17.9 15.4 23.2









#40
Hexachlorocyclopentadiene
Concen: 35.177 ng/ul
RT: 12.669 min Scan# 1629
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SLCS690


Tgt Ion:237 Resp: 248447
Ion Ratio Lower Upper
237 100
235 64.0 50.5 75.7
272 14.2 10.5 15.7



#41
2,4,6-Trichlorophenol
Concen: 41.133 ng/ul
RT: 12.916 min Scan# 1671
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48

Tgt Ion:196 Resp: 245153
Ion Ratio Lower Upper
196 100
198 96.2 74.9 112.3
200 28.2 23.8 35.8









#42
2,4,5-Trichlorophenol
Concen: 40.699 ng/ul
RT: 12.987 min Scan# 1683
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SLCS690


Tgt Ion:196 Resp: 281447
Ion Ratio Lower Upper
196 100
198 96.5 76.4 114.6
200 31.8 24.7 37.1



#43
1,1'-Biphenyl
Concen: 38.951 ng/ul
RT: 13.328 min Scan# 1741
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48

Tgt Ion:154 Resp: 1059483
Ion Ratio Lower Upper
154 100
153 40.5 32.1 48.1
76 11.8 9.8 14.6









#44
2-Chloronaphthalene
Concen: 39.031 ng/ul
RT: 13.363 min Scan# 1747
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SLCS690


Tgt Ion:162 Resp: 835788
Ion Ratio Lower Upper
162 100
164 33.4 25.4 38.0
127 39.2 30.5 45.7



#45
2-Nitroaniline
Concen: 39.896 ng/ul
RT: 13.557 min Scan# 1780
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48


Tgt Ion: 65 Resp: 194563
Ion Ratio Lower Upper
65 100
92 67.8 56.6 84.8
138 110.9 90.6 135.8








#47
Dimethylphthalate
Concen: 39.953 ng/ul
RT: 13.951 min Scan# 1847
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48

Instrument :
BNA_N
ClientSampleId :
SLCS690







Tgt Ion:163 Resp: 942164
Ion Ratio Lower Upper
163 100
194 4.5 3.5 5.3
164 10.1 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 41.098 ng/ul
RT: 14.057 min Scan# 1865
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48

Tgt Ion:165 Resp: 173279
Ion Ratio Lower Upper
165 100
89 57.8 48.2 72.2
121 22.7 18.2 27.2









#50
Acenaphthylene
Concen: 39.781 ng/ul
RT: 14.216 min Scan# 1892
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SLCS690


Tgt Ion:152 Resp: 1360428
Ion Ratio Lower Upper
152 100
151 19.1 15.7 23.5
153 12.0 10.1 15.1



#51
3-Nitroaniline
Concen: 42.046 ng/ul
RT: 14.387 min Scan# 1921
Delta R.T. 0.006 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48

Tgt Ion:138 Resp: 189997
Ion Ratio Lower Upper
138 100
108 12.0 9.9 14.9
92 108.2 91.0 136.4





#52

Acenaphthene

Concen: 39.602 ng/ul

RT: 14.557 min Scan# 1950

Delta R.T. -0.000 min

Lab File: BN037732.D

Acq: 16 Sep 2025 14:48

Instrument :

BNA_N

ClientSampleId :

SLCS690





Tgt Ion:153 Resp: 881281

Ion Ratio Lower Upper

153 100

152 47.8 38.4 57.6

154 87.9 73.1 109.7



#53

2,4-Dinitrophenol

Concen: 32.832 ng/ul

RT: 14.587 min Scan# 1955

Delta R.T. -0.000 min

Lab File: BN037732.D

Acq: 16 Sep 2025 14:48




Tgt Ion:184 Resp: 58506


Ion Ratio Lower Upper

184 100

63 66.8 52.5 78.7

154 70.1 57.6 86.4









#55
4-Nitrophenol
Concen: 40.145 ng/ul
RT: 14.687 min Scan# 1971
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SLCS690


Tgt Ion:109 Resp: 135118
Ion Ratio Lower Upper
109 100
139 141.5 113.7 170.5
65 123.0 104.5 156.7



#56
Dibenzofuran
Concen: 39.242 ng/ul
RT: 14.892 min Scan# 2007
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48

Tgt Ion:168 Resp: 1191439
Ion Ratio Lower Upper
168 100
139 38.3 30.2 45.2









#57
2,4-Dinitrotoluene
Concen: 43.530 ng/ul
RT: 14.845 min Scan# 1999
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SLCS690


Tgt Ion:165 Resp: 262783
Ion Ratio Lower Upper
165 100
63 41.5 34.2 51.2
182 3.1 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 44.211 ng/ul
RT: 15.116 min Scan# 2045
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48

Tgt Ion:232 Resp: 213828
Ion Ratio Lower Upper
232 100
131 47.8 37.5 56.3
130 2.5 1.8 2.8
166 32.4 25.4 38.0





#59

Diethylphthalate

Concen: 40.453 ng/ul

RT: 15.322 min Scan# 2080

Delta R.T. -0.000 min

Lab File: BN037732.D

Acq: 16 Sep 2025 14:48

Instrument :

BNA_N

ClientSampleId :

SLCS690





Tgt Ion:149 Resp: 906406

Ion Ratio Lower Upper

149 100

177 21.4 18.1 27.1

150 11.6 9.8 14.8



#61

Fluorene

Concen: 39.725 ng/ul

RT: 15.545 min Scan# 2118

Delta R.T. -0.000 min

Lab File: BN037732.D

Acq: 16 Sep 2025 14:48




Tgt Ion:166 Resp: 977499


Ion Ratio Lower Upper

166 100

165 93.7 79.4 119.2

167 13.7 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 39.094 ng/ul
RT: 15.539 min Scan# 2117
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SLCS690


Tgt Ion:204 Resp: 462848
Ion Ratio Lower Upper
204 100
206 34.6 25.9 38.9
141 60.3 46.0 69.0



#63
4-Nitroaniline
Concen: 51.851 ng/ul
RT: 15.545 min Scan# 2118
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48

Tgt Ion:138 Resp: 220438
Ion Ratio Lower Upper
138 100
92 48.2 39.2 58.8
108 75.6 57.0 85.4









#66
 4,6-Dinitro-2-methylphenol
 Concen: 34.938 ng/ul
 RT: 15.610 min Scan# 2129
 Delta R.T. -0.000 min
 Lab File: BN037732.D
 Acq: 16 Sep 2025 14:48




Instrument :
 BNA_N
 ClientSampleId :
 SLCS690


Tgt Ion:198 Resp: 112042
 Ion Ratio Lower Upper
 198 100
 182 4.9 3.7 5.5
 77 24.5 20.2 30.4



#67
 N-Nitrosodiphenylamine
 Concen: 39.762 ng/ul
 RT: 15.751 min Scan# 2153
 Delta R.T. -0.000 min
 Lab File: BN037732.D
 Acq: 16 Sep 2025 14:48

Tgt Ion:169 Resp: 812094
 Ion Ratio Lower Upper
 169 100
 168 66.2 53.9 80.9
 167 36.1 29.2 43.8









#68
4-Bromophenyl-phenylether
Concen: 40.511 ng/ul
RT: 16.433 min Scan# 211
Delta R.T. -0.006 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SLCS690


Tgt Ion:248 Resp: 290112
Ion Ratio Lower Upper
248 100
250 94.8 73.6 110.4
141 73.7 51.8 77.8



#69
Hexachlorobenzene
Concen: 39.509 ng/ul
RT: 16.551 min Scan# 2289
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48

Tgt Ion:284 Resp: 346296
Ion Ratio Lower Upper
284 100
142 32.3 25.5 38.3
249 29.9 23.9 35.9









#70
Atrazine
Concen: 37.265 ng/ul
RT: 16.704 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SLCS690


Tgt Ion:200 Resp: 255750
Ion Ratio Lower Upper
200 100
173 26.5 21.9 32.9
215 48.3 39.0 58.4



#71
Pentachlorophenol
Concen: 38.289 ng/ul
RT: 16.886 min Scan# 2346
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48

Tgt Ion:266 Resp: 184227
Ion Ratio Lower Upper
266 100
264 64.3 48.3 72.5
268 63.8 48.3 72.5









#72
 Phenanthrene
 Concen: 39.666 ng/u1
 RT: 17.281 min Scan# 2413
 Delta R.T. -0.006 min
 Lab File: BN037732.D
 Acq: 16 Sep 2025 14:48




Instrument :
 BNA_N
 ClientSampleId :
 SLCS690


Tgt Ion:178 Resp: 1463612
 Ion Ratio Lower Upper
 178 100
 179 15.1 12.4 18.6
 176 18.0 15.3 22.9



#74
 Anthracene
 Concen: 39.890 ng/u1
 RT: 17.375 min Scan# 2429
 Delta R.T. -0.000 min
 Lab File: BN037732.D
 Acq: 16 Sep 2025 14:48

Tgt Ion:178 Resp: 1492446
 Ion Ratio Lower Upper
 178 100
 179 16.3 12.3 18.5
 176 19.0 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 36.581 ng/uL
RT: 13.292 min Scan# 1735
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SLCS690


Tgt Ion:216 Resp: 414426
Ion Ratio Lower Upper
216 100
214 76.8 64.0 96.0
218 46.8 39.6 59.4



#76
Pentachlorobenzene
Concen: 37.515 ng/uL
RT: 14.816 min Scan# 1994
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48

Tgt Ion:250 Resp: 407113
Ion Ratio Lower Upper
250 100
248 64.9 50.6 75.8
252 65.7 50.9 76.3









#77
 Carbazole
 Concen: 40.197 ng/ul
 RT: 17.639 min Scan# 2474
 Delta R.T. -0.000 min
 Lab File: BN037732.D
 Acq: 16 Sep 2025 14:48




Instrument :
 BNA_N
 ClientSampleId :
 SLCS690


Tgt Ion:167 Resp: 1332321
 Ion Ratio Lower Upper
 167 100
 166 21.4 16.6 24.8
 139 12.9 10.1 15.1



#78
 Di-n-butylphthalate
 Concen: 43.362 ng/ul
 RT: 18.222 min Scan# 2573
 Delta R.T. -0.000 min
 Lab File: BN037732.D
 Acq: 16 Sep 2025 14:48

Tgt Ion:149 Resp: 1458400
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.3 10.9
 104 5.0 4.0 6.0





#80

Fluoranthene

Concen: 39.920 ng/ul

RT: 19.298 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN037732.D

Acq: 16 Sep 2025 14:48

Instrument :

BNA_N

ClientSampleId :

SLCS690





Tgt Ion:202 Resp: 1605399

Ion Ratio Lower Upper

202 100

101 10.8 9.0 13.4

100 8.2 7.0 10.6



#82

Pyrene

Concen: 39.686 ng/ul

RT: 19.657 min Scan# 2817

Delta R.T. -0.000 min

Lab File: BN037732.D

Acq: 16 Sep 2025 14:48




Tgt Ion:202 Resp: 1697432


Ion Ratio Lower Upper

202 100

101 13.1 10.9 16.3

100 11.4 9.0 13.6





#83

Butylbenzylphthalate

Concen: 44.640 ng/ul

RT: 20.563 min Scan# 2971

Delta R.T. -0.006 min

Lab File: BN037732.D

Acq: 16 Sep 2025 14:48

Instrument :

BNA_N

ClientSampleId :

SLCS690





Tgt Ion:149 Resp: 539050

Ion Ratio Lower Upper

149 100

91 70.9 53.1 79.7

206 21.2 17.1 25.7



#84

3,3'-Dichlorobenzidine

Concen: 37.664 ng/ul

RT: 21.392 min Scan# 3112

Delta R.T. -0.000 min

Lab File: BN037732.D

Acq: 16 Sep 2025 14:48




Tgt Ion:252 Resp: 479575


Ion Ratio Lower Upper

252 100

254 67.6 50.2 75.2

126 11.6 9.2 13.8









#85
Benzo(a)anthracene
Concen: 39.181 ng/ul
RT: 21.469 min Scan# 3126
Delta R.T. -0.006 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SLCS690


Tgt Ion:228 Resp: 1656640
Ion Ratio Lower Upper
228 100
229 21.3 15.4 23.0
226 25.0 21.9 32.9



#86
Bis(2-ethylhexyl)phthalate
Concen: 45.785 ng/ul
RT: 21.416 min Scan# 3116
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48


Tgt Ion:149 Resp: 833440
Ion Ratio Lower Upper
149 100
167 26.9 21.3 31.9
279 5.0 3.4 5.0








#87
Chrysene
Concen: 39.166 ng/ul
RT: 21.533 min Scan# 3136
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48

Instrument :
BNA_N
ClientSampleId :
SLCS690







Tgt Ion:228 Resp: 1583444
Ion Ratio Lower Upper
228 100
226 30.8 24.1 36.1
229 20.0 16.8 25.2



#89
Di-n-octyl phthalate
Concen: 37.959 ng/ul
RT: 22.586 min Scan# 3315
Delta R.T. -0.006 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48

Tgt Ion:149 Resp: 1239740





#90

Benzo(b)fluoranthene

Concen: 41.416 ng/ul

RT: 23.598 min Scan# 3487

Delta R.T. -0.006 min

Lab File: BN037732.D

Acq: 16 Sep 2025 14:48

Instrument :

BNA_N

ClientSampleId :

SLCS690





Tgt Ion:252 Resp: 1695413

Ion Ratio Lower Upper

252 100

253 21.6 16.6 24.8

125 9.7 8.1 12.1



#91

Benzo(k)fluoranthene

Concen: 40.980 ng/ul

RT: 23.668 min Scan# 3499

Delta R.T. -0.000 min

Lab File: BN037732.D

Acq: 16 Sep 2025 14:48




Tgt Ion:252 Resp: 1692754


Ion Ratio Lower Upper

252 100

253 22.9 16.9 25.3

125 10.4 8.1 12.1









#93
Benzo(a)pyrene
Concen: 41.778 ng/ul
RT: 24.433 min Scan# 3629
Delta R.T. -0.000 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SLCS690


Tgt Ion:252 Resp: 1639290
Ion Ratio Lower Upper
252 100
253 20.8 17.3 25.9
125 11.3 8.9 13.3



#94
Indeno(1,2,3-cd)pyrene
Concen: 44.009 ng/ul
RT: 28.068 min Scan# 4247
Delta R.T. 0.006 min
Lab File: BN037732.D
Acq: 16 Sep 2025 14:48

Tgt Ion:276 Resp: 2259547
Ion Ratio Lower Upper
276 100
138 18.2 15.3 22.9
277 22.2 18.0 27.0









#95
 Dibenzo(a,h)anthracene
 Concen: 44.173 ng/ul
 RT: 28.139 min Scan# 41
 Delta R.T. 0.006 min
 Lab File: BN037732.D
 Acq: 16 Sep 2025 14:48

Instrument :
 BNA_N
 ClientSampleId :
 SLCS690

Tgt Ion:278 Resp: 1874585

Ion	Ratio	Lower	Upper
278	100		
139	15.2	12.6	19.0
279	23.0	19.4	29.0



#96
 Benzo(g,h,i)perylene
 Concen: 43.691 ng/ul
 RT: 29.150 min Scan# 4431
 Delta R.T. -0.006 min
 Lab File: BN037732.D
 Acq: 16 Sep 2025 14:48

Tgt Ion:276 Resp: 1838111

Ion	Ratio	Lower	Upper
276	100		
138	20.3	16.7	25.1
277	23.5	18.8	28.2

