

FORM 8A-OR
INTERNAL STANDARD AREA AND RETENTION TIME SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: ARC Sayre MA No.: 30184323.0001 SDG No.: Q3084
 Analytical Method: Trace VOA Level : _____
 EPA Sample No. : VSTD005358 Lab File ID (Standard) : VV039119.D
 Instrument ID : MSVOA_V Init.Calib.Date(s): 09/15/2025
 GC Column : DB-624UI ID: 0.18 (mm) Date Analyzed : 09/17/2025
 Heated Purge: (Y/N) N Time Analyzed : 09:29

	IS1 (DFB) AREA	RT	IS2 (CBZ) AREA	RT	IS3 (DCB) AREA	RT
12 HOUR STD	381055	5.53	388618	8.78	191056	11.17
UPPER LIMIT	762110	5.702	777236	8.946	382112	11.342
LOWER LIMIT	190528	5.362	194309	8.606	95528	11.002
EPA SAMPLE NO.						
VBLK291	355233	5.54	369520	8.78	131601	11.18
VHBLK001	321294	5.54	329760	8.78	123971	11.18

IS1 (DFB) = 1,4-Difluorobenzene
 IS2 (CBZ) = Chlorobenzene-d5
 IS3 (DCB) = 1,4-Dichlorobenzene -d4

AREA UPPER LIMIT = 200% of internal standard area
 AREA LOWER LIMIT = 50% of internal standard area
 RT UPPER LIMIT = +0.17 minutes of internal standard RT
 RT LOWER LIMIT = -0.17 minutes of internal standard RT