

## LAB CHRONICLE

|                 |                                                   |                   |                              |
|-----------------|---------------------------------------------------|-------------------|------------------------------|
| <b>OrderID:</b> | Q3085                                             | <b>OrderDate:</b> | 9/12/2025 10:08:00 AM        |
| <b>Client:</b>  | Core Environmental Consultants and Services, Inc. | <b>Project:</b>   | Bayshore Soil BSM Management |
| <b>Contact:</b> | Ronald Tramosch                                   | <b>Location:</b>  | D31                          |

| LabID    | ClientID     | Matrix | Test              | Method | Sample Date | Prep Date | Anal Date | Received |
|----------|--------------|--------|-------------------|--------|-------------|-----------|-----------|----------|
| Q3085-01 | PAULDING-WC1 | TCLP   | TCLPMetals Group3 | 6010D  | 08/29/25    | 09/15/25  | 09/15/25  | 08/29/25 |



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**Hit Summary Sheet**  
**SW-846**

|                 |                                                   |                    |                              |
|-----------------|---------------------------------------------------|--------------------|------------------------------|
| <b>SDG No.:</b> | Q3085                                             | <b>Order ID:</b>   | Q3085                        |
| <b>Client:</b>  | Core Environmental Consultants and Services, Inc. | <b>Project ID:</b> | Bayshore Soil BSM Management |

| Sample ID               | Client ID                    | Matrix | Parameter | Concentration | C | MDL  | RDL  | Units |
|-------------------------|------------------------------|--------|-----------|---------------|---|------|------|-------|
| Client ID :<br>Q3085-01 | PAULDING-WC1<br>PAULDING-WC1 | TCLP   | Lead      | 36.2          | J | 11.5 | 60.0 | ug/L  |



# SAMPLE DATA

## Report of Analysis

|                   |                                                   |                 |          |
|-------------------|---------------------------------------------------|-----------------|----------|
| Client:           | Core Environmental Consultants and Services, Inc. | Date Collected: | 08/29/25 |
| Project:          | Bayshore Soil BSM Management                      | Date Received:  | 08/29/25 |
| Client Sample ID: | PAULDING-WC1                                      | SDG No.:        | Q3085    |
| Lab Sample ID:    | Q3085-01                                          | Matrix:         | TCLP     |
| Level (low/med):  | low                                               | % Solid:        | 0        |

| Cas       | Parameter | Conc. | Qua. | DF | MDL  | LOQ / CRQL | Units | Prep Date      | Date Ana.      | Ana Met. | Prep Met. |
|-----------|-----------|-------|------|----|------|------------|-------|----------------|----------------|----------|-----------|
| 7439-92-1 | Lead      | 36.2  | J    | 1  | 11.5 | 60.0       | ug/L  | 09/15/25 12:30 | 09/15/25 15:49 | 6010D    | SW3050    |

|               |                    |                 |       |            |
|---------------|--------------------|-----------------|-------|------------|
| Color Before: | Colorless          | Clarity Before: | Clear | Texture:   |
| Color After:  | Colorless          | Clarity After:  | Clear | Artifacts: |
| Comments:     | TCLP Metals Group3 |                 |       |            |

U = Not Detected  
 LOQ = Limit of Quantitation  
 MDL = Method Detection Limit  
 LOD = Limit of Detection  
 D = Dilution  
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value  
 B = Analyte Found in Associated Method Blank  
 \* = indicates the duplicate analysis is not within control limits.  
 E = Indicates the reported value is estimated because of the presence of interference.  
 OR = Over Range  
 N = Spiked sample recovery not within control limits



# METAL CALIBRATION DATA

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** Core Environmental Consultants and Services, In

**SDG No.:** Q3085

**Contract:** CORE02

**Lab Code:** ACE

**Initial Calibration Source:** EPA

**Continuing Calibration Source:** Inorganic Ventures

| Sample ID | Analyte | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| ICV01     | Lead    | 3730           | 4000       | 93            | 90 - 110                  | P | 09/15/2025       | 11:31            | LB137198      |

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** Core Environmental Consultants and Services, Inc.

**SDG No.:** Q3085

**Contract:** CORE02

**Lab Code:** ACE

**Initial Calibration Source:** EPA

**Continuing Calibration Source:** Inorganic Ventures

| Sample ID | Analyte | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| LLICV01   | Lead    | 11.7           | 12.0       | 98            | 80 - 120                  | P | 09/15/2025       | 11:36            | LB137198      |

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** Core Environmental Consultants and Services, Inc.

**SDG No.:** Q3085

**Contract:** CORE02

**Lab Code:** ACE

**Initial Calibration Source:** EPA

**Continuing Calibration Source:** Inorganic Ventures

| Sample ID | Analyte | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| CCV01     | Lead    | 4940           | 5000       | 99            | 90 - 110                  | P | 09/15/2025       | 12:35            | LB137198      |
| CCV02     | Lead    | 4710           | 5000       | 94            | 90 - 110                  | P | 09/15/2025       | 13:42            | LB137198      |
| CCV03     | Lead    | 4810           | 5000       | 96            | 90 - 110                  | P | 09/15/2025       | 15:11            | LB137198      |
| CCV04     | Lead    | 4840           | 5000       | 97            | 90 - 110                  | P | 09/15/2025       | 16:12            | LB137198      |
| CCV05     | Lead    | 4760           | 5000       | 95            | 90 - 110                  | P | 09/15/2025       | 16:55            | LB137198      |



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**Metals**

**- 2b -**

**CRDL STANDARD FOR AA & ICP**

**Client:** Core Environmental Consultants and Services, Inc.

**SDG No.:** Q3085

**Contract:** CORE02

**Lab Code:** ACE

**Initial Calibration Source:** \_\_\_\_\_

**Continuing Calibration Source:** \_\_\_\_\_

| Sample ID | Analyte | Result<br>ug/L | True Value<br>ug/L | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|--------------------|---------------|---------------------------|---|------------------|------------------|---------------|
| CRI01     | Lead    | 11.4           | 12.0               | 95            | 65 - 135                  | P | 09/15/2025       | 11:47            | LB137198      |



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**Metals**

**- 3a -**

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** Core Environmental Consultants and Services, Inc.

**SDG No.:** Q3085

**Contract:** CORE02

**Lab Code:** ACE

| Sample ID | Analyte | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| ICB01     | Lead    | 2.30           | +/-6                | U            | 12.0 | P | 09/15/2025       | 11:43            | LB137198      |

**Metals**

- 3a -

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** Core Environmental Consultants and Services, Inc.

**SDG No.:** Q3085

**Contract:** CORE02

**Lab Code:** ACE

| Sample ID | Analyte | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| CCB01     | Lead    | 2.30           | +/-6                | U            | 12.0 | P | 09/15/2025       | 12:39            | LB137198      |
| CCB02     | Lead    | 2.30           | +/-6                | U            | 12.0 | P | 09/15/2025       | 13:47            | LB137198      |
| CCB03     | Lead    | 2.30           | +/-6                | U            | 12.0 | P | 09/15/2025       | 15:15            | LB137198      |
| CCB04     | Lead    | 2.30           | +/-6                | U            | 12.0 | P | 09/15/2025       | 16:16            | LB137198      |
| CCB05     | Lead    | 2.30           | +/-6                | U            | 12.0 | P | 09/15/2025       | 16:59            | LB137198      |

**Metals**  
**- 3b -**  
**PREPARATION BLANK SUMMARY**

**Client:** Core Environmental Consultants and Services, **SDG No.:** Q3085

**Instrument:** P4

| Sample ID         | Analyte | Result<br>(ug/L) | Acceptance<br>Limit | Conc<br>Qual         | CRQL<br>ug/L    | M | Analysis<br>Date  | Analysis<br>Time  | Run      |
|-------------------|---------|------------------|---------------------|----------------------|-----------------|---|-------------------|-------------------|----------|
| <b>PB169677TB</b> |         | <b>WATER</b>     |                     | <b>Batch Number:</b> | <b>PB169685</b> |   | <b>Prep Date:</b> | <b>09/15/2025</b> |          |
|                   | Lead    | 11.5             | <30                 | U                    | 60.0            | P | 09/15/2025        | 16:50             | LB137198 |
| Sample ID         | Analyte | Result<br>(ug/L) | Acceptance<br>Limit | Conc<br>Qual         | CRQL<br>ug/L    | M | Analysis<br>Date  | Analysis<br>Time  | Run      |
| <b>PB169685BL</b> |         | <b>WATER</b>     |                     | <b>Batch Number:</b> | <b>PB169685</b> |   | <b>Prep Date:</b> | <b>09/15/2025</b> |          |
|                   | Lead    | 11.5             | <30                 | U                    | 60.0            | P | 09/15/2025        | 15:19             | LB137198 |

**Metals**

- 4 -

**INTERFERENCE CHECK SAMPLE**

**Client:** Core Environmental Consultants and Services, Inc.

**SDG No.:** Q3085

**Contract:** CORE02

**Lab Code:** ACE

**ICS Source:** EPA

**Instrument ID:** P4

| Sample ID | Analyte | Result<br>ug/L | True Value<br>ug/L | %<br>Recovery | Low<br>Limit<br>(ug/L) | High<br>Limit<br>(ug/L) | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|--------------------|---------------|------------------------|-------------------------|------------------|------------------|---------------|
| ICSA01    | Lead    | -0.061         |                    |               | -12                    | 12                      | 09/15/2025       | 12:12            | LB137198      |
| ICSAB01   | Lead    | 45.1           | 49.0               | 92            | 37                     | 61                      | 09/15/2025       | 12:16            | LB137198      |



# METAL QC DATA

metals  
- 5a -  
MATRIX SPIKE SUMMARY

**client:** Core Environmental Consultants and Service    **level:** low    **sdg no.:** Q3085  
**contract:** CORE02    **lab code:** ACE  
**matrix:** Water    **sample id:** Q3092-04    **client id:** 279182MS  
**Percent Solids for Sample:** NA    **Spiked ID:** Q3092-04MS    **Percent Solids for Spike Sample:** NA

| Analyte | Units | Acceptance<br>Limit %R | Spiked<br>Result | C | Sample<br>Result | C | Spike<br>Added | %<br>Recovery | Qual | M |
|---------|-------|------------------------|------------------|---|------------------|---|----------------|---------------|------|---|
| Lead    | ug/L  | 75 - 125               | 4150             |   | 19.3             | J | 5000           | 83            |      | P |

metals  
- 5a -  
MATRIX SPIKE DUPLICATE SUMMARY

client: Core Environmental Consultants and Service level: low sdg no.: Q3085  
contract: CORE02 lab code: ACE  
matrix: Water sample id: Q3092-04 client id: 279182MSD  
Percent Solids for Sample: NA Spiked ID: Q3092-04MSD Percent Solids for Spike Sample: NA

| Analyte | Units | Acceptance<br>Limit %R | MSD<br>Result | C | Sample<br>Result | C | Spike<br>Added | %<br>Recovery | Qual | M |
|---------|-------|------------------------|---------------|---|------------------|---|----------------|---------------|------|---|
| Lead    | ug/L  | 75 - 125               | 4510          |   | 19.3             | J | 5000           | 90            |      | P |

**Metals**  
**- 5b -**

**Client:** Core Environmental Consultants and Servic

**SDG No.:** Q3085

**Contract:** CORE02

**Lab Code:** ACE

**Matrix:**

**Level:** LOW

**Client ID:**

**Sample ID:**  **Spiked ID:**

| Analyte | Units | Acceptance<br>Limit %R | C | Sample<br>Result | C | Spike<br>Added | %<br>Recovery | Qual | M |
|---------|-------|------------------------|---|------------------|---|----------------|---------------|------|---|
|---------|-------|------------------------|---|------------------|---|----------------|---------------|------|---|

A  
B  
C  
D  
E  
F  
G  
H

**Metals**

- 6 -

**DUPLICATE SAMPLE SUMMARY**

**Client:** Core Environmental Consultants and Service **Level:** LOW **SDG No.:** Q3085  
**Contract:** CORE02 **Lab Code:** ACE  
**Matrix:** Water **Sample ID:** Q3092-04 **Client ID:** 279182DUP  
**Percent Solids for Sample:** NA **Duplicate ID** Q3092-04DUP **Percent Solids for Spike Sample:** NA

| Analyte | Units | Acceptance Limit | Sample Result | C | Duplicate Result | C | RPD | Qual | M |
|---------|-------|------------------|---------------|---|------------------|---|-----|------|---|
| Lead    | ug/L  | 20               | 19.3          | J | 16.2             | J | 17  |      | P |

**Metals**

- 6 -

**DUPLICATE SAMPLE SUMMARY**

**Client:** Core Environmental Consultants and Service    **Level:** LOW    **SDG No.:** Q3085  
**Contract:** CORE02    **Lab Code:** ACE  
**Matrix:** Water    **Sample ID:** Q3092-04MS    **Client ID:** 279182MSD  
**Percent Solids for Sample:** NA    **Duplicate ID** Q3092-04MSD    **Percent Solids for Spike Sample:** NA

| Analyte | Units | Acceptance Limit | Sample Result | C | Duplicate Result | C | RPD | Qual | M |
|---------|-------|------------------|---------------|---|------------------|---|-----|------|---|
| Lead    | ug/L  | 20               | 4150          |   | 4510             |   | 8   |      | P |

**Metals**

- 7 -

**LABORATORY CONTROL SAMPLE SUMMARY**

**Client:** Core Environmental Consultants and Service **SDG No.:** Q3085  
**Contract:** CORE02 **Lab Code:** ACE

| Analyte            | Units | True Value | Result | C | %<br>Recovery | Acceptance<br>Limits | M |
|--------------------|-------|------------|--------|---|---------------|----------------------|---|
| PB169685BS<br>Lead | ug/L  | 5000       | 4640   |   | 93            | 80 - 120             | P |

**Metals**  
**-9 -**  
**ICP SERIAL DILUTIONS**

**SAMPLE NO.**

279182L

**Lab Name:** Alliance **Contract:** CORE02  
**Lab Code:** ACE **Lb No.:** lb137198 **Lab Sample ID :** Q3092-04L **SDG No.:** Q3085  
**Matrix (soil/water):** Water **Level (low/med):** LOW  
**Concentration Units:** ug/L

| Analyte | Initial Sample<br>Result (I) |   | Serial Dilution<br>Result (S) |   | % Differ-<br>ence | Q | M |
|---------|------------------------------|---|-------------------------------|---|-------------------|---|---|
| Lead    | 19.3                         | J | 300                           | U | 100.0             |   | P |



# METAL PREPARATION & INSTRUMENT DATA

**Metals**

- 11 -

**ICP INTERELEMENT CORRECTION FACTORS**

**Client:** Core Environmental Consultants and Service

**SDG No.:** Q3085

**Contract:** CORE02

**Lab Code:** ACE

**Instrument ID:** \_\_\_\_\_

**Date:** \_\_\_\_\_

**Interelement Correction Factors (apparent ppb analyte/ppm interferent )**

| Analyte | Wave-<br>Length<br>(nm) | ICP Interelement Correction Factors For: |           |           |           |           |
|---------|-------------------------|------------------------------------------|-----------|-----------|-----------|-----------|
|         |                         | Al                                       | Ca        | Fe        | Mg        | Ag        |
| Lead    | 220.353                 | -0.0000920                               | 0.0000000 | 0.0000380 | 0.0000000 | 0.0000000 |

**Metals**

- 11 -

**ICP INTERELEMENT CORRECTION FACTORS**

**Client:** Core Environmental Consultants and Service

**SDG No.:** Q3085

**Contract:** CORE02

**Lab Code:** ACE

**Instrument ID:** \_\_\_\_\_

**Date:** \_\_\_\_\_

**Interelement Correction Factors (apparent ppb analyte/ppm interferent )**

| Analyte | Wave-<br>Length<br>(nm) | ICP Interelement Correction Factors For: |           |           |           |           |
|---------|-------------------------|------------------------------------------|-----------|-----------|-----------|-----------|
|         |                         | As                                       | Ba        | Be        | Cd        | Co        |
| Lead    | 220.353                 | 0.0000000                                | 0.0003170 | 0.0000000 | 0.0000000 | 0.0000000 |

**Metals**

- 11 -

**ICP INTERELEMENT CORRECTION FACTORS**

**Client:** Core Environmental Consultants and Service

**SDG No.:** Q3085

**Contract:** CORE02

**Lab Code:** ACE

**Instrument ID:** \_\_\_\_\_

**Date:** \_\_\_\_\_

**Interelement Correction Factors (apparent ppb analyte/ppm interferent )**

| Analyte | Wave-<br>Length<br>(nm) | ICP Interelement Correction Factors For: |           |           |           |            |
|---------|-------------------------|------------------------------------------|-----------|-----------|-----------|------------|
|         |                         | Cr                                       | Cu        | K         | Mn        | Mo         |
| Lead    | 220.353                 | 0.0000000                                | 0.0000000 | 0.0000000 | 0.0001400 | -0.0008600 |

**Metals**

- 11 -

**ICP INTERELEMENT CORRECTION FACTORS**

**Client:** Core Environmental Consultants and Service

**SDG No.:** Q3085

**Contract:** CORE02

**Lab Code:** ACE

**Instrument ID:** \_\_\_\_\_

**Date:** \_\_\_\_\_

**Interelement Correction Factors (apparent ppb analyte/ppm interferent )**

| Analyte | Wave-<br>Length<br>(nm) | ICP Interelement Correction Factors For: |           |           |           |           |
|---------|-------------------------|------------------------------------------|-----------|-----------|-----------|-----------|
|         |                         | Na                                       | Ni        | Pb        | Sb        | Se        |
| Lead    | 220.353                 | 0.0000000                                | 0.0006580 | 0.0000000 | 0.0000000 | 0.0001290 |

**Metals**

- 11 -

**ICP INTERELEMENT CORRECTION FACTORS**

**Client:** Core Environmental Consultants and Service

**SDG No.:** Q3085

**Contract:** CORE02

**Lab Code:** ACE

**Instrument ID:** \_\_\_\_\_

**Date:** \_\_\_\_\_

**Interelement Correction Factors (apparent ppb analyte/ppm interferent )**

| Analyte | Wave-<br>Length<br>(nm) | ICP Interelement Correction Factors For: |            |           |           |           |
|---------|-------------------------|------------------------------------------|------------|-----------|-----------|-----------|
|         |                         | Sn                                       | Ti         | Tl        | V         | Zn        |
| Lead    | 220.353                 | 0.0000000                                | -0.0003610 | 0.0000000 | 0.0000000 | 0.0000000 |

# METAL PREPARATION & ANALYICAL SUMMARY

**Metals**

- 13 -

**SAMPLE PREPARATION SUMMARY**

**Client:** Core Environmental Consultants and Service **SDG No.:** Q3085  
**Contract:** CORE02 **Lab Code:** ACE **Method:** \_\_\_\_\_

| Sample ID                     | Client ID    | Sample Type | Matrix | Prep Date  | Initial Sample Size(mL) | Final Sample Volume (mL) | Percent Solids |
|-------------------------------|--------------|-------------|--------|------------|-------------------------|--------------------------|----------------|
| <b>Batch Number: PB169685</b> |              |             |        |            |                         |                          |                |
| PB169677TB                    | PB169677TB   | MB          | WATER  | 09/15/2025 | 5.0                     | 25.0                     |                |
| PB169685BL                    | PB169685BL   | MB          | WATER  | 09/15/2025 | 5.0                     | 25.0                     |                |
| PB169685BS                    | PB169685BS   | LCS         | WATER  | 09/15/2025 | 5.0                     | 25.0                     |                |
| Q3085-01                      | PAULDING-WC1 | SAM         | WATER  | 09/15/2025 | 5.0                     | 25.0                     |                |
| Q3092-04DUP                   | 279182DUP    | DUP         | WATER  | 09/15/2025 | 5.0                     | 25.0                     |                |
| Q3092-04MS                    | 279182MS     | MS          | WATER  | 09/15/2025 | 5.0                     | 25.0                     |                |
| Q3092-04MSD                   | 279182MSD    | MSD         | WATER  | 09/15/2025 | 5.0                     | 25.0                     |                |

**metals**  
**- 14 -**  
**ANALYSIS RUN LOG**

**Client:** Core Environmental Consultants and Service

**Contract:** CORE02

**Lab code:** ACE

**Sdg no.:** Q3085

**Instrument id number:** \_\_\_\_\_

**Method:** \_\_\_\_\_

**Run number:** LB137198

**Start date:** 09/15/2025

**End date:** 09/15/2025

| Lab sample id. | Client Sample Id | d/f | Time | Parameter list |
|----------------|------------------|-----|------|----------------|
| S0             | S0               | 1   | 1106 | Pb             |
| S1             | S1               | 1   | 1110 | Pb             |
| S2             | S2               | 1   | 1114 | Pb             |
| S3             | S3               | 1   | 1118 | Pb             |
| S4             | S4               | 1   | 1122 | Pb             |
| S5             | S5               | 1   | 1126 | Pb             |
| ICV01          | ICV01            | 1   | 1131 | Pb             |
| LLICV01        | LLICV01          | 1   | 1136 | Pb             |
| ICB01          | ICB01            | 1   | 1143 | Pb             |
| CRI01          | CRI01            | 1   | 1147 | Pb             |
| ICSA01         | ICSA01           | 1   | 1212 | Pb             |
| ICSAB01        | ICSAB01          | 1   | 1216 | Pb             |
| CCV01          | CCV01            | 1   | 1235 | Pb             |
| CCB01          | CCB01            | 1   | 1239 | Pb             |
| CCV02          | CCV02            | 1   | 1342 | Pb             |
| CCB02          | CCB02            | 1   | 1347 | Pb             |
| CCV03          | CCV03            | 1   | 1511 | Pb             |
| CCB03          | CCB03            | 1   | 1515 | Pb             |
| PB169685BL     | PB169685BL       | 1   | 1519 | Pb             |
| PB169685BS     | PB169685BS       | 1   | 1523 | Pb             |
| Q3085-01       | PAULDING-WC1     | 1   | 1549 | Pb             |
| CCV04          | CCV04            | 1   | 1612 | Pb             |
| CCB04          | CCB04            | 1   | 1616 | Pb             |
| Q3092-04DUP    | 279182DUP        | 1   | 1629 | Pb             |
| Q3092-04L      | 279182L          | 5   | 1634 | Pb             |
| Q3092-04MS     | 279182MS         | 1   | 1638 | Pb             |
| Q3092-04MSD    | 279182MSD        | 1   | 1642 | Pb             |
| PB169677TB     | PB169677TB       | 1   | 1650 | Pb             |
| CCV05          | CCV05            | 1   | 1655 | Pb             |
| CCB05          | CCB05            | 1   | 1659 | Pb             |